

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: YANN01
 Lab Code: CHEM Case No.: P4473 SAS No.: P4473 SDG NO.: P4473
 Lab File ID: VY020001.D Date Analyzed: 10/24/2024
 Instrument ID: MSVOA_Y Time Analyzed: 10:18
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	173628	13.346				
UPPER LIMIT	347256	13.846				
LOWER LIMIT	86814	12.846				
EPA SAMPLE NO.						
TS-1	98425	13.35				
VY1024SBL01	125898	13.35				
VY1024SBS01	144922	13.35				
VY1024SBSD01	155771	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.