



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Yannuzzi Group, Inc.

SDG No.: P4474

Project: 86 Davidson Road, Piscataway, NJ

RunNo.: LB133078

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.095	0.099	96	90-110	10/23/2024
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	10/23/2024
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	10/23/2024