



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: Yannuzzi Group, Inc.

SDG No.: P4474

Project: 86 Davidson Road, Piscataway, NJ

RunNo.: LB133078

Analyte		Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID:	ICB1							
Cyanide		mg/L	0.001	0.0025	J	0.00099	0.005	10/23/2024
Sample ID:	CCB1							
Cyanide		mg/L	0.0011	0.0025	J	0.00099	0.005	10/23/2024
Sample ID:	CCB2							
Cyanide		mg/L	0.001	0.0025	J	0.00099	0.005	10/23/2024