

Duplicate Sample Summary

Client:	Yannuzzi Group, Inc.	SDG No.:	P4474
Project:	86 Davidson Road, Piscataway, NJ	Sample ID:	P4473-01
Client ID:	TS-1DUP	Percent Solids for Spike Sample:	48.8

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	0.78		0.76		1	3		10/23/2024

Duplicate Sample Summary

Client:	Yannuzzi Group, Inc.	SDG No.:	P4474
Project:	86 Davidson Road, Piscataway, NJ	Sample ID:	P4473-01
Client ID:	TS-1MSD	Percent Solids for Spike Sample:	48.8

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	4.80		4.90		1	2		10/23/2024