

CALIBRATION VERIFICATION SUMMARY

Contract: YANN01

Lab Code: CHEM Case No.: P4474 SAS No.: P4474 SDG NO.: P4474

Continuing Calib Date: 10/24/2024 Initial Calibration Date(s): 10/21/2024 10/21/2024

Continuing Calib Time: 17:20 Initial Calibration Time(s): 13:00 13:53

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.55	3.54	3.44	3.64	-0.01
alpha-BHC	4.00	3.99	3.89	4.09	-0.01
beta-BHC	4.53	4.52	4.42	4.62	-0.01
delta-BHC	4.78	4.77	4.67	4.87	-0.01
gamma-BHC (Lindane)	4.34	4.33	4.23	4.43	-0.01
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.69	5.68	5.58	5.78	-0.01
Endosulfan I	6.08	6.07	5.97	6.17	-0.01
Dieldrin	6.35	6.35	6.25	6.45	0.00
4,4'-DDE	6.20	6.19	6.09	6.29	-0.01
Endrin	6.58	6.57	6.47	6.67	-0.01
Endosulfan II	6.80	6.79	6.69	6.89	-0.01
4,4'-DDD	6.72	6.71	6.61	6.81	-0.01
Endosulfan sulfate	7.17	7.16	7.06	7.26	-0.01
4,4'-DDT	7.03	7.02	6.92	7.12	-0.01
Methoxychlor	7.51	7.50	7.40	7.60	-0.01
Endrin ketone	7.65	7.64	7.54	7.74	-0.01
Endrin aldehyde	6.93	6.92	6.82	7.02	-0.01
alpha-Chlordane	6.03	6.02	5.92	6.12	-0.01
gamma-Chlordane	5.95	5.94	5.84	6.04	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: YANN01

Lab Code: CHEM Case No.: P4474 SAS No.: P4474 SDG NO.: P4474

Continuing Calib Date: 10/24/2024 Initial Calibration Date(s): 10/21/2024 10/21/2024

Continuing Calib Time: 17:20 Initial Calibration Time(s): 13:00 13:53

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.92	7.92	7.82	8.02	0.00
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00
alpha-BHC	3.28	3.28	3.18	3.38	0.00
beta-BHC	3.91	3.91	3.81	4.01	0.00
delta-BHC	4.14	4.14	4.04	4.24	0.00
gamma-BHC (Lindane)	3.61	3.61	3.51	3.71	0.00
Heptachlor	3.95	3.95	3.85	4.05	0.00
Aldrin	4.23	4.23	4.13	4.33	0.00
Heptachlor epoxide	4.74	4.73	4.63	4.83	-0.01
Endosulfan I	5.11	5.10	5.00	5.20	-0.01
Dieldrin	5.37	5.37	5.27	5.47	0.00
4,4'-DDE	5.24	5.23	5.13	5.33	-0.01
Endrin	5.65	5.64	5.54	5.74	0.00
Endosulfan II	5.94	5.94	5.84	6.04	0.00
4,4'-DDD	5.79	5.79	5.69	5.89	0.00
Endosulfan sulfate	6.34	6.34	6.24	6.44	0.00
4,4'-DDT	6.04	6.04	5.94	6.14	0.00
Methoxychlor	6.62	6.62	6.52	6.72	0.00
Endrin ketone	6.85	6.84	6.74	6.94	-0.01
Endrin aldehyde	6.12	6.12	6.02	6.22	0.00
alpha-Chlordane	5.05	5.05	4.95	5.15	0.00
gamma-Chlordane	4.99	4.98	4.88	5.08	-0.01