

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4474 SAS No.: P4474 SDG NO.: P4474

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/24/2024

Lab File ID: BF140001.D Time Analyzed: 15:16

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	551060	11.416	264546	14.051	308388	15.533
UPPER LIMIT	1102120	11.916	529092	14.551	616776	16.033
LOWER LIMIT	275530	10.916	132273	13.551	154194	15.033
EPA SAMPLE NO.						
01 TP-1MS	529411	11.42	267551	14.05	360724	15.53
02 TP-1MSD	450738	11.42	233244	14.05	308426	15.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.