

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4475 SAS No.: P4475 SDG NO.: P4475

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 10/24/2024

Lab File ID: BN034686.D Time Analyzed: 10:23

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	6324	7.712	18497	10.48	9311	14.33
UPPER LIMIT	12648	8.212	36994	10.979	18622	14.829
LOWER LIMIT	3162	7.212	9248.5	9.979	4655.5	13.829
EPA SAMPLE NO.						
01 BP-VPB-190-GW-238-240	6066	7.71	17751	10.48	8148	14.33
02 BP-VPB-190-GW-278-280	5957	7.71	17174	10.48	7846	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.