

Surrogate Summary

SDG No.: P4475

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4475-01	BP-VPB-190-TB-20241018	1,2-Dichloroethane-d4	50	46.0	92	81	118
		Dibromofluoromethane	50	50.8	102	80	119
		Toluene-d8	50	49.1	98	89	112
		4-Bromofluorobenzene	50	45.6	91	85	114
P4475-02	BP-VPB-190-DUP-20241018	1,2-Dichloroethane-d4	50	49.6	99	81	118
		Dibromofluoromethane	50	52.6	105	80	119
		Toluene-d8	50	50.4	101	89	112
		4-Bromofluorobenzene	50	47.5	95	85	114
P4475-03	BP-VPB-190-GW-223-225	1,2-Dichloroethane-d4	50	48.5	97	81	118
		Dibromofluoromethane	50	52.3	105	80	119
		Toluene-d8	50	49.8	99	89	112
		4-Bromofluorobenzene	50	46.1	92	85	114
P4475-04	BP-VPB-190-GW-238-240	1,2-Dichloroethane-d4	50	46.3	93	81	118
		Dibromofluoromethane	50	50.2	100	80	119
		Toluene-d8	50	48.9	98	89	112
		4-Bromofluorobenzene	50	46.4	93	85	114
P4475-05	BP-VPB-190-GW-258-260	1,2-Dichloroethane-d4	50	49.0	98	81	118
		Dibromofluoromethane	50	51.3	103	80	119
		Toluene-d8	50	49.9	100	89	112
		4-Bromofluorobenzene	50	47.6	95	85	114
P4475-06	BP-VPB-190-GW-278-280	1,2-Dichloroethane-d4	50	50.0	100	81	118
		Dibromofluoromethane	50	51.3	103	80	119
		Toluene-d8	50	49.9	100	89	112
		4-Bromofluorobenzene	50	46.4	93	85	114
VN1022WBL01	VN1022WBL01	1,2-Dichloroethane-d4	50	45.7	91	81	118
		Dibromofluoromethane	50	50.0	100	80	119
		Toluene-d8	50	49.3	99	89	112
		4-Bromofluorobenzene	50	45.0	90	85	114
VN1022WBS01	VN1022WBS01	1,2-Dichloroethane-d4	50	46.2	92	81	118
		Dibromofluoromethane	50	49.5	99	80	119
		Toluene-d8	50	48.8	98	89	112
		4-Bromofluorobenzene	50	49.0	98	85	114
VN1022WBSD01	VN1022WBSD01	1,2-Dichloroethane-d4	50	49.1	98	81	118
		Dibromofluoromethane	50	49.5	99	80	119
		Toluene-d8	50	48.0	96	89	112
		4-Bromofluorobenzene	50	48.9	98	85	114