

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE450
 Preparation Blank Matrix : Soil
 Preparation Blank Concentration Units (µg/L, mg/L, mg/kg dry weight, or µg): mg/kg
 Analytical Method: ICP-MS Preparation Batch: PB164345
 Run Batch: LB133210 Preparation Method: 200.8

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID: ICB007	Q	ID: CCB035	Q	ID: CCB036	Q	ID: CCB037	Q	ID: PBS345	Q
Antimony	2.0	U	0.35	J	0.28	J	0.34	J	1.0	U
Arsenic	1.0	U	1.0	U	1.0	U	1.0	U	0.5	U
Barium	10.0	U	10.0	U	10.0	U	10.0	U	5.0	U
Beryllium	1.0	U	1.0	U	1.0	U	1.0	U	0.5	U
Cadmium	1.0	U	1.0	U	1.0	U	1.0	U	0.5	U
Chromium	2.0	U	2.0	U	2.0	U	2.0	U	1.0	U
Cobalt	1.0	U	1.0	U	1.0	U	1.0	U	0.5	U
Copper	2.0	U	0.26	J	0.29	J	0.42	J	1.0	U
Lead	1.0	U	0.19	J	0.16	J	0.24	J	0.5	U
Nickel	1.0	U	1.0	U	1.0	U	1.0	U	0.5	U
Selenium	5.0	U	5.0	U	5.0	U	5.0	U	2.5	U
Silver	1.0	U	1.0	U	1.0	U	1.0	U	0.5	U
Thallium	1.0	U	0.08	J	0.08	J	0.09	J	0.5	U
Vanadium	5.0	U	5.0	U	5.0	U	0.04	J	2.5	U
Zinc	5.0	U	5.0	U	5.0	U	5.0	U	2.5	U

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE450
Preparation Blank Matrix :
Preparation Blank Concentration Units ($\mu\text{g/L}$, mg/L , mg/kg dry weight, or μg):
Analytical Method: ICP-MS Preparation Batch:
Run Batch: LB133210 Preparation Method:

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID:	Q	ID: CCB038	Q	ID: CCB039	Q	ID: CCB040	Q	ID:	Q
Antimony			0.32	J	0.3	J	0.35	J		
Arsenic			1.0	U	1.0	U	1.0	U		
Barium			10.0	U	10.0	U	10.0	U		
Beryllium			1.0	U	1.0	U	1.0	U		
Cadmium			1.0	U	1.0	U	1.0	U		
Chromium			2.0	U	2.0	U	2.0	U		
Cobalt			1.0	U	1.0	U	1.0	U		
Copper			2.0	U	2.0	U	0.35	J		
Lead			0.18	J	0.16	J	0.27	J		
Nickel			1.0	U	1.0	U	0.23	J		
Selenium			5.0	U	5.0	U	5.0	U		
Silver			1.0	U	1.0	U	1.0	U		
Thallium			0.08	J	1.0	U	0.09	J		
Vanadium			5.0	U	5.0	U	0.05	J		
Zinc			5.0	U	5.0	U	5.0	U		