

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE490Initial Calibration Verification Source : MP82450Continuing Calibration Verification Source : MP82453Run Batch: LB133197 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV004				ID: CCV023				ID: CCV024		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2300	92	1	400000	410000	103	0	420000	104	0
Antimony	1000	1000	100	0	5000	4700	95	2	4800	95	0
Arsenic	1000	1100	105	1	5000	5000	100	2	5000	99	0
Barium	520	500	96	0	10000	10000	100	0	10000	101	1
Beryllium	510	520	101	1	500	490	99	0	510	102	0
Cadmium	510	490	96	0	2500	2500	99	2	2500	99	0
Calcium	10000	9900	99	0	400000	400000	101	0	410000	103	0
Chromium	520	550	107	0	15500	15000	100	0	16000	100	0
Cobalt	520	490	93	0	2500	2500	100	2	2500	100	0
Copper	510	480	95	0	15000	15000	100	0	15000	101	0
Iron	10000	11000	106	1	400000	410000	103	0	420000	105	0
Lead	1000	960	96	0	25000	24000	98	2	24000	98	0
Magnesium	6000	5700	94	1	400000	400000	101	0	410000	102	0
Manganese	520	530	101	0	15000	15000	99	0	15000	99	1
Nickel	530	500	94	0	2500	2500	99	2	2500	99	0
Potassium	9900	9100	92	0	150000	150000	102	0	160000	104	0
Selenium	1000	1000	103	1	5000	4600	93	2	4700	93	0
Silver	250	240	95	0	1250	1200	99	0	1300	102	0
Sodium	10000	9500	95	3	400000	410000	102	0	410000	104	0
Thallium	1000	1000	103	1	5000	4800	97	2	4800	96	0
Vanadium	500	520	103	1	2500	2500	99	1	2500	101	1
Zinc	1000	1000	100	0	15000	14000	93	1	14000	94	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE490Initial Calibration Verification Source : MP82450Continuing Calibration Verification Source : MP82453Run Batch: LB133197 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV025				ID: CCV026		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	420000	104	1	420000	104	0
Antimony					5000	4700	95	1	4800	95	1
Arsenic					5000	5000	100	0	5000	100	0
Barium					10000	10000	101	0	10000	102	0
Beryllium					500	510	102	0	510	101	0
Cadmium					2500	2500	99	0	2500	99	0
Calcium					400000	410000	102	1	410000	102	0
Chromium					15500	15000	100	0	16000	100	0
Cobalt					2500	2500	100	0	2500	100	0
Copper					15000	15000	100	0	15000	101	0
Iron					400000	410000	103	1	410000	103	0
Lead					25000	24000	98	0	24000	98	0
Magnesium					400000	400000	101	1	400000	101	0
Manganese					15000	15000	99	1	15000	99	0
Nickel					2500	2500	99	0	2500	99	1
Potassium					150000	160000	104	0	160000	104	0
Selenium					5000	4700	93	0	4700	94	1
Silver					1250	1300	101	0	1300	101	0
Sodium					400000	410000	103	0	420000	104	0
Thallium					5000	4800	96	0	4800	97	1
Vanadium					2500	2500	100	1	2500	100	1
Zinc					15000	14000	94	0	14000	94	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE490Initial Calibration Verification Source : MP83143Continuing Calibration Verification Source : MP83146Run Batch: LB133346 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV005				ID: CCV029				ID: CCV030		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2300	93	0	400000	400000	100	0	400000	100	3
Antimony	1000	900	90	1	5000	4700	95	0	4800	96	0
Arsenic	1000	1000	102	1	5000	4900	98	0	5000	100	0
Barium	520	500	97	0	10000	10000	100	1	9900	99	3
Beryllium	510	520	102	0	500	500	100	0	500	100	2
Cadmium	510	490	95	0	2500	2500	98	0	2500	99	0
Calcium	10000	10000	100	1	400000	400000	100	0	400000	100	3
Chromium	520	550	106	0	15500	15000	99	0	15000	97	3
Cobalt	520	490	93	0	2500	2500	99	0	2500	100	0
Copper	510	480	95	1	15000	15000	99	0	15000	97	3
Iron	10000	11000	106	0	400000	410000	102	0	410000	102	3
Lead	1000	960	96	0	25000	24000	97	0	25000	98	0
Magnesium	6000	6000	99	1	400000	390000	97	0	390000	96	3
Manganese	520	520	100	1	15000	15000	98	0	14000	96	3
Nickel	530	500	94	1	2500	2500	99	0	2500	99	0
Potassium	9900	9500	96	1	150000	150000	100	0	150000	100	3
Selenium	1000	930	93	2	5000	4700	94	0	4700	95	0
Silver	250	240	95	0	1250	1200	99	0	1200	99	3
Sodium	10000	9600	96	1	400000	400000	101	0	400000	101	3
Thallium	1000	1000	103	1	5000	4800	96	1	4800	96	0
Vanadium	500	520	103	2	2500	2500	98	1	2400	98	3
Zinc	1000	930	93	0	15000	14000	96	1	14000	95	4

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE490Initial Calibration Verification Source : MP83143Continuing Calibration Verification Source : MP83146Run Batch: LB133346 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV031				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	101	0			
Antimony					5000	4800	95	0			
Arsenic					5000	5000	100	1			
Barium					10000	10000	101	0			
Beryllium					500	510	102	0			
Cadmium					2500	2500	99	0			
Calcium					400000	400000	101	0			
Chromium					15500	15000	98	0			
Cobalt					2500	2500	99	0			
Copper					15000	15000	99	0			
Iron					400000	410000	103	0			
Lead					25000	24000	98	0			
Magnesium					400000	390000	97	1			
Manganese					15000	15000	98	0			
Nickel					2500	2500	99	0			
Potassium					150000	150000	102	0			
Selenium					5000	4700	94	0			
Silver					1250	1200	100	0			
Sodium					400000	410000	102	0			
Thallium					5000	4800	95	0			
Vanadium					2500	2500	99	1			
Zinc					15000	14000	96	0			