

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE007	TUNE	10/30/24 11:23		Jaswal	OK
2	S00	S0	CAL1	10/30/24 12:18		Jaswal	OK
3	S02	S02	CAL3	10/30/24 12:25		Jaswal	OK
4	S03	S03	CAL4	10/30/24 12:28		Jaswal	OK
5	S04	S04	CAL5	10/30/24 12:31		Jaswal	OK
6	S05	S05	CAL6	10/30/24 12:34		Jaswal	OK
7	S06	S06	CAL7	10/30/24 12:37		Jaswal	OK
8	S07	S07	CAL8	10/30/24 12:40		Jaswal	OK
9	S08	S08	CAL9	10/30/24 12:43		Jaswal	OK
10	ICV007	ICV007	ICV	10/30/24 12:59		Jaswal	OK
11	ICB007	ICB007	ICB	10/30/24 13:04		Jaswal	OK
12	ICSA007	ICSA007	ICSA	10/30/24 13:07		Jaswal	OK
13	ICSAB007	ICSAB007	ICSAB	10/30/24 13:10		Jaswal	OK
14	CCV035	CCV035	CCV	10/30/24 13:13		Jaswal	OK
15	CCB035	CCB035	CCB	10/30/24 13:16		Jaswal	OK
16	PB164277BL	PBS277	MB	10/30/24 13:19		Jaswal	OK
17	PB164277BS	LCS277	LCS	10/30/24 13:23		Jaswal	OK
18	PB164277BSD	LCSD277	LCSD	10/30/24 13:26		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

19	P4331-04DL	MYDBB7	SAM	10/30/24 13:29	4X For Ba,Co,Ni,V	Jaswal	Confirms
20	P4331-05DL	MYDBB7D	DUP	10/30/24 13:33	4X For Ba,Co,Ni,V	Jaswal	Confirms
21	P4331-04LDL	MYDBB7L	SD	10/30/24 13:36	4X For Ba,Co,Ni,V	Jaswal	Confirms
22	P4331-06REDL	MYDBB7S	MS	10/30/24 13:39	Not use	Jaswal	Not Ok
23	P4331-04REMSDL	MYDBB7SRE	MS	10/30/24 13:42	Not use	Jaswal	Not Ok
24	P4331-06DL	MYDBB7S	MS	10/30/24 13:48	4X For Ba,Co,Ni,V	Jaswal	Confirms
25	P4331-04MSDL	MYDBB7SRE	MS	10/30/24 13:51	4X For Ba,Co,Ni,V	Jaswal	Confirms
26	P4331-04A	MYDBB7ARE	PS	10/30/24 13:54	PS for Ag-0.082ml(MP82908)-10 ml	Jaswal	OK
27	CCV036	CCV036	CCV	10/30/24 13:57		Jaswal	OK
28	CCB036	CCB036	CCB	10/30/24 14:00		Jaswal	OK
29	PB164345BL	PBS345	MB	10/30/24 14:05		Jaswal	OK
30	PB164345BSD	LCSD345	LCSD	10/30/24 14:12		Jaswal	OK
31	PB164349BL	PBS349	MB	10/30/24 14:15		Jaswal	OK
32	PB164349BS	LCS349	LCS	10/30/24 14:24		Jaswal	OK
33	PB164349BSD	LCSD349	LCSD	10/30/24 14:28		Jaswal	OK
34	PB164427BL	PBS427	MB	10/30/24 14:31		Jaswal	OK
35	PB164427BS	LCS427	LCS	10/30/24 14:34		Jaswal	OK
36	PB164427BSD	LCSD427	LCSD	10/30/24 14:37		Jaswal	OK
37	PB164345BS	LCS345	LCS	10/30/24 14:41		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

38	P4477-04	MYE453	SAM	10/30/24 14:44	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
39	P4477-05	MYE453D	DUP	10/30/24 14:47	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
40	P4477-04L	MYE453L	SD	10/30/24 14:50	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
41	P4477-06	MYE453S	MS	10/30/24 14:54	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
42	P4477-04MS	MYE453SRE	MS	10/30/24 14:57	MS Fail for Ag, Ba,Co,Ni,V High	Jaswal	Dilution
43	P4479-12	MYE481	SAM	10/30/24 15:00	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
44	P4479-13	MYE481D	DUP	10/30/24 15:03	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
45	P4479-12L	MYE481L	SD	10/30/24 15:06	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
46	P4479-14	MYE481S	MS	10/30/24 15:09	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
47	P4479-12MS	MYE481SRE	MS	10/30/24 15:12	MS Fail for Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
48	P4480-05	MYE494	SAM	10/30/24 15:15	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
49	P4480-06	MYE494D	DUP	10/30/24 15:19	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
50	P4480-05L	MYE494L	SD	10/30/24 15:22	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
51	P4480-07	MYE494S	MS	10/30/24 15:25	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
52	CCV037	CCV037	CCV	10/30/24 15:28		Jaswal	OK
53	CCB037	CCB037	CCB	10/30/24 15:31		Jaswal	OK
54	P4480-05MS	MYE494SRE	MS	10/30/24 15:34	MS Fail for Ag(Below RL),MS Fail for Ag, Ba,Co,Ni,V High	Jaswal	Dilution
55	P4477-01	MYE450	SAM	10/30/24 15:37		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

56	P4477-02	MYE451	SAM	10/30/24 15:41		Jaswal	OK
57	P4477-03	MYE452	SAM	10/30/24 15:44		Jaswal	OK
58	P4477-07	MYE454	SAM	10/30/24 15:47		Jaswal	OK
59	P4477-08	MYE455	SAM	10/30/24 15:50		Jaswal	OK
60	P4477-09	MYE456	SAM	10/30/24 15:53		Jaswal	OK
61	P4477-10	MYE457	SAM	10/30/24 15:56		Jaswal	OK
62	P4477-11	MYE458	SAM	10/30/24 16:00		Jaswal	OK
63	P4477-12	MYE459	SAM	10/30/24 16:03		Jaswal	OK
64	P4477-13	MYE460	SAM	10/30/24 16:06		Jaswal	OK
65	P4477-14	MYE461	SAM	10/30/24 16:09		Jaswal	OK
66	P4477-15	MYE462	SAM	10/30/24 16:12		Jaswal	OK
67	P4477-16	MYE463	SAM	10/30/24 16:15		Jaswal	OK
68	P4477-17	MYE464	SAM	10/30/24 16:19		Jaswal	OK
69	P4477-18	MYE465	SAM	10/30/24 16:22		Jaswal	OK
70	P4477-19	MYE466	SAM	10/30/24 16:25		Jaswal	OK
71	P4477-20	MYE467	SAM	10/30/24 16:28		Jaswal	OK
72	P4477-21	MYE468	SAM	10/30/24 16:31		Jaswal	OK
73	P4477-22	MYE469	SAM	10/30/24 16:34		Jaswal	OK
74	P4479-01	MYE470	SAM	10/30/24 16:38		Jaswal	OK
75	P4479-02	MYE471	SAM	10/30/24 16:41		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

76	P4479-03	MYE472	SAM	10/30/24 16:44		Jaswal	OK
77	P4479-04	MYE473	SAM	10/30/24 16:47		Jaswal	OK
78	P4479-05	MYE474	SAM	10/30/24 16:50		Jaswal	OK
79	P4479-06	MYE475	SAM	10/30/24 16:53		Jaswal	OK
80	CCV038	CCV038	CCV	10/30/24 16:57		Jaswal	OK
81	CCB038	CCB038	CCB	10/30/24 16:59		Jaswal	OK
82	P4479-07	MYE476	SAM	10/30/24 17:03		Jaswal	OK
83	P4479-08	MYE477	SAM	10/30/24 17:06		Jaswal	OK
84	P4479-09	MYE478	SAM	10/30/24 17:09		Jaswal	OK
85	P4479-10	MYE479	SAM	10/30/24 17:12		Jaswal	OK
86	P4479-11	MYE480	SAM	10/30/24 17:15		Jaswal	OK
87	P4479-15	MYE482	SAM	10/30/24 17:19		Jaswal	OK
88	P4479-16	MYE483	SAM	10/30/24 17:22		Jaswal	OK
89	P4479-17	MYE484	SAM	10/30/24 17:25		Jaswal	OK
90	P4479-18	MYE485	SAM	10/30/24 17:28		Jaswal	OK
91	P4479-19	MYE486	SAM	10/30/24 17:31		Jaswal	OK
92	P4479-20	MYE487	SAM	10/30/24 17:34		Jaswal	OK
93	P4479-21	MYE488	SAM	10/30/24 17:38		Jaswal	OK
94	P4479-22	MYE489	SAM	10/30/24 17:41		Jaswal	OK
95	P4480-01	MYE490	SAM	10/30/24 17:44		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

96	P4480-02	MYE491	SAM	10/30/24 17:47		Jaswal	OK
97	P4480-03	MYE492	SAM	10/30/24 17:50		Jaswal	OK
98	P4480-04	MYE493	SAM	10/30/24 17:53		Jaswal	OK
99	P4480-08	MYE495	SAM	10/30/24 17:56		Jaswal	OK
100	P4480-09	MYE496	SAM	10/30/24 18:00		Jaswal	OK
101	P4480-10	MYE497	SAM	10/30/24 18:03		Jaswal	OK
102	P4480-11	MYE498	SAM	10/30/24 18:06		Jaswal	OK
103	P4480-12	MYE499	SAM	10/30/24 18:09		Jaswal	OK
104	P4480-13	MYE4A0	SAM	10/30/24 18:12		Jaswal	OK
105	P4480-14	MYE4A1	SAM	10/30/24 18:16		Jaswal	OK
106	P4480-15	MYE4A2	SAM	10/30/24 18:19		Jaswal	OK
107	P4480-16	MYE4A3	SAM	10/30/24 18:22		Jaswal	OK
108	CCV039	CCV039	CCV	10/30/24 18:25		Jaswal	OK
109	CCB039	CCB039	CCB	10/30/24 18:28		Jaswal	OK
110	P4480-17	MYE4A4	SAM	10/30/24 18:31		Jaswal	OK
111	P4480-18	MYE4A5	SAM	10/30/24 18:34		Jaswal	OK
112	P4480-19	MYE4A6	SAM	10/30/24 18:37		Jaswal	OK
113	P4480-20	MYE4A7	SAM	10/30/24 18:41		Jaswal	OK
114	P4480-21	MYE4A8	SAM	10/30/24 18:44		Jaswal	OK
115	P4480-22	MYE4A9	SAM	10/30/24 18:47		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

116	P4477-04DL	MYE453	SAM	10/30/24 18:50	4X For Ba,Co,Ni,V High	Jaswal	Confirms
117	P4477-05DL	MYE453D	DUP	10/30/24 18:53	4X For Ba,Co,Ni,V High	Jaswal	Confirms
118	P4477-04LDL	MYE453L	SD	10/30/24 18:56	20X For Ba,Co,Ni,V High	Jaswal	Confirms
119	P4477-06DL	MYE453S	MS	10/30/24 19:00	4X For Ba,Co,Ni,V High	Jaswal	Confirms
120	P4477-04MSDL	MYE453SRE	MS	10/30/24 19:03	4X For Ba,Co,Ni,V High	Jaswal	Confirms
121	P4479-12DL	MYE481	SAM	10/30/24 19:06	4X For Ba,Co,Ni,V High	Jaswal	Confirms
122	P4479-13DL	MYE481D	DUP	10/30/24 19:09	4X For Ba,Co,Ni,V High	Jaswal	Confirms
123	P4479-12LDL	MYE481L	SD	10/30/24 19:13	20X For Ba,Co,Ni,V High	Jaswal	Confirms
124	P4479-14DL	MYE481S	MS	10/30/24 19:16	4X For Ba,Co,Ni,V High	Jaswal	Confirms
125	P4479-12MSDL	MYE481SRE	MS	10/30/24 19:19	4X For Ba,Co,Ni,V High	Jaswal	Confirms
126	P4480-05DL	MYE494	SAM	10/30/24 19:22	4X For Ba,Co,Ni,V High	Jaswal	Confirms
127	P4480-06DL	MYE494D	DUP	10/30/24 19:25	4X For Ba,Co,Ni,V High	Jaswal	Confirms
128	P4480-05LDL	MYE494L	SD	10/30/24 19:29	20X For Ba,Co,Ni,V High	Jaswal	Confirms
129	P4480-07DL	MYE494S	MS	10/30/24 19:32	4X For Ba,Co,Ni,V High	Jaswal	Confirms
130	P4480-05MSDL	MYE494SRE	MS	10/30/24 19:35	4X For Ba,Co,Ni,V High	Jaswal	Confirms
131	P4477-04A	MYE453ARE	PS	10/30/24 19:38	PS For AG(0.02ML MP82908-10ML SAMPLE)	Jaswal	OK
132	P4479-12A	MYE481ARE	PS	10/30/24 19:41	PS For AG(0.02ML MP82908-10ML SAMPLE)	Jaswal	OK
133	P4480-05A	MYE494ARE	PS	10/30/24 19:45	PS For AG(0.02ML MP82908-10ML SAMPLE)	Jaswal	OK
134	PB164430BL	PBS430	MB	10/30/24 19:54		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

135	PB164430BS	LCS430	LCS	10/30/24 19:57		Jaswal	OK
136	CCV040	CCV040	CCV	10/30/24 20:01		Jaswal	OK
137	CCB040	CCB040	CCB	10/30/24 20:03		Jaswal	OK
138	PB164430BSD	LCSD430	LCSD	10/30/24 20:07		Jaswal	OK
139	PB164440BL	PBS440	MB	10/30/24 20:10		Jaswal	OK
140	PB164440BS	LCS440	LCS	10/30/24 20:13		Jaswal	OK
141	PB164440BSD	LCSD440	LCSD	10/30/24 20:17		Jaswal	OK
142	P4481-12	MYE4C1	SAM	10/30/24 20:20	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
143	P4481-13	MYE4C1D	DUP	10/30/24 20:23	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
144	P4481-12L	MYE4C1L	SD	10/30/24 20:26	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
145	P4481-14	MYE4C1S	MS	10/30/24 20:29	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
146	P4481-12MS	MYE4C1SRE	MS	10/30/24 20:33	Ba,Co,Ni,V High	Jaswal	Dilution
147	P4482-14	MYE4E3	SAM	10/30/24 20:36	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
148	P4482-15	MYE4E3D	DUP	10/30/24 20:39	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
149	P4482-14L	MYE4E3L	SD	10/30/24 20:42	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
150	P4482-16	MYE4E3S	MS	10/30/24 20:45	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
151	P4482-14MS	MYE4E3SRE	MS	10/30/24 20:48	Ba,Co,Ni,V High	Jaswal	Dilution
152	P4483-20	MYE4G9	SAM	10/30/24 20:51	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
153	P4483-21	MYE4G9D	DUP	10/30/24 20:55	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
154	P4483-20L	MYE4G9L	SD	10/30/24 20:58	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

155	P4483-22	MYE4G9S	MS	10/30/24 21:01	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
156	P4483-20MS	MYE4G9SRE	MS	10/30/24 21:04	MS Fail for Pb,Ag(Below rl),Ba,Co,Ni,V High	Jaswal	Dilution
157	PB164442BL	PBS442	MB	10/30/24 21:07		Jaswal	OK
158	PB164442BS	LCS442	LCS	10/30/24 21:10		Jaswal	OK
159	PB164442BSD	LCSD442	LCSD	10/30/24 21:14		Jaswal	OK
160	P4481-01	MYE4B0	SAM	10/30/24 21:17		Jaswal	OK
161	P4481-02	MYE4B1	SAM	10/30/24 21:20		Jaswal	OK
162	P4481-03	MYE4B2	SAM	10/30/24 21:23		Jaswal	OK
163	P4481-04	MYE4B3	SAM	10/30/24 21:27		Jaswal	OK
164	CCV041	CCV041	CCV	10/30/24 21:30		Jaswal	OK
165	CCB041	CCB041	CCB	10/30/24 21:33		Jaswal	OK
166	P4481-05	MYE4B4	SAM	10/30/24 21:36		Jaswal	OK
167	P4481-06	MYE4B5	SAM	10/30/24 21:39		Jaswal	OK
168	P4481-07	MYE4B6	SAM	10/30/24 21:42		Jaswal	OK
169	P4481-08	MYE4B7	SAM	10/30/24 21:45		Jaswal	OK
170	P4481-09	MYE4B8	SAM	10/30/24 21:49		Jaswal	OK
171	P4481-10	MYE4B9	SAM	10/30/24 21:52		Jaswal	OK
172	P4481-11	MYE4C0	SAM	10/30/24 21:55		Jaswal	OK
173	P4481-15	MYE4C2	SAM	10/30/24 21:58		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

174	P4481-16	MYE4C3	SAM	10/30/24 22:01		Jaswal	OK
175	P4481-17	MYE4C4	SAM	10/30/24 22:05		Jaswal	OK
176	P4481-18	MYE4C5	SAM	10/30/24 22:08		Jaswal	OK
177	P4481-19	MYE4C6	SAM	10/30/24 22:11		Jaswal	OK
178	P4481-20	MYE4C7	SAM	10/30/24 22:14		Jaswal	OK
179	P4481-21	MYE4C8	SAM	10/30/24 22:17		Jaswal	OK
180	P4481-22	MYE4C9	SAM	10/30/24 22:21		Jaswal	OK
181	P4482-01	MYE4D0	SAM	10/30/24 22:24		Jaswal	OK
182	P4482-02	MYE4D1	SAM	10/30/24 22:27		Jaswal	OK
183	P4482-03	MYE4D2	SAM	10/30/24 22:30		Jaswal	OK
184	P4482-04	MYE4D3	SAM	10/30/24 22:33		Jaswal	OK
185	P4482-05	MYE4D4	SAM	10/30/24 22:36		Jaswal	OK
186	P4482-06	MYE4D5	SAM	10/30/24 22:40		Jaswal	OK
187	P4482-07	MYE4D6	SAM	10/30/24 22:43		Jaswal	OK
188	P4482-08	MYE4D7	SAM	10/30/24 22:46		Jaswal	OK
189	P4482-09	MYE4D8	SAM	10/30/24 22:50		Jaswal	OK
190	P4482-10	MYE4D9	SAM	10/30/24 22:53		Jaswal	OK
191	P4482-11	MYE4E0	SAM	10/30/24 22:57		Jaswal	OK
192	CCV042	CCV042	CCV	10/30/24 23:00		Jaswal	OK
193	CCB042	CCB042	CCB	10/30/24 23:03		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

194	P4482-12	MYE4E1	SAM	10/30/24 23:06		Jaswal	OK
195	P4482-13	MYE4E2	SAM	10/30/24 23:09		Jaswal	OK
196	P4482-17	MYE4E4	SAM	10/30/24 23:12		Jaswal	OK
197	P4482-18	MYE4E5	SAM	10/30/24 23:15		Jaswal	OK
198	P4482-19	MYE4E6	SAM	10/30/24 23:19		Jaswal	OK
199	P4482-20	MYE4E7	SAM	10/30/24 23:22		Jaswal	OK
200	P4482-21	MYE4E8	SAM	10/30/24 23:25		Jaswal	OK
201	P4482-22	MYE4E9	SAM	10/30/24 23:28		Jaswal	OK
202	P4483-01	MYE4F0	SAM	10/30/24 23:31		Jaswal	OK
203	P4483-02	MYE4F1	SAM	10/30/24 23:35		Jaswal	OK
204	P4483-03	MYE4F2	SAM	10/30/24 23:38		Jaswal	OK
205	P4483-04	MYE4F3	SAM	10/30/24 23:41		Jaswal	OK
206	P4483-05	MYE4F4	SAM	10/30/24 23:44		Jaswal	OK
207	P4483-06	MYE4F5	SAM	10/30/24 23:47		Jaswal	OK
208	P4483-07	MYE4F6	SAM	10/30/24 23:50		Jaswal	OK
209	P4483-08	MYE4F7	SAM	10/30/24 23:54		Jaswal	OK
210	P4483-09	MYE4F8	SAM	10/30/24 23:57		Jaswal	OK
211	P4483-10	MYE4F9	SAM	10/31/24		Jaswal	OK
212	P4483-11	MYE4G0	SAM	10/31/24 00:03		Jaswal	OK
213	ICSA008	ICSA008	ICSA	10/31/24 00:06		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

214	ICSAB008	ICSAB008	ICSAB	10/31/24 00:10		Jaswal	OK
215	CCV043	CCV043	CCV	10/31/24 00:13		Jaswal	OK
216	CCB043	CCB043	CCB	10/31/24 00:16		Jaswal	OK
217	P4483-12	MYE4G1	SAM	10/31/24 00:19		Jaswal	OK
218	P4483-13	MYE4G2	SAM	10/31/24 00:22		Jaswal	OK
219	P4483-14	MYE4G3	SAM	10/31/24 00:25		Jaswal	OK
220	P4483-15	MYE4G4	SAM	10/31/24 00:28		Jaswal	OK
221	P4483-16	MYE4G5	SAM	10/31/24 00:32		Jaswal	OK
222	P4483-17	MYE4G6	SAM	10/31/24 00:35		Jaswal	OK
223	P4483-18	MYE4G7	SAM	10/31/24 00:38		Jaswal	OK
224	P4483-19	MYE4G8	SAM	10/31/24 00:41		Jaswal	OK
225	P4481-12DL	MYE4C1	SAM	10/31/24 00:44	4X For Ba,Co,Ni,V	Jaswal	Confirms
226	P4481-13DL	MYE4C1D	DUP	10/31/24 00:48	4X For Ba,Co,Ni,V	Jaswal	Confirms
227	P4481-12LDL	MYE4C1L	SD	10/31/24 00:51	20X For Ba,Co,Ni,V	Jaswal	Confirms
228	P4481-14DL	MYE4C1S	MS	10/31/24 00:54	4X For Ba,Co,Ni,V	Jaswal	Confirms
229	P4481-12MSDL	MYE4C1SRE	MS	10/31/24 00:57	4X For Ba,Co,Ni,V	Jaswal	Confirms
230	P4482-14DL	MYE4E3	SAM	10/31/24 01:01	4X For Ba,Co,Ni,V	Jaswal	Confirms
231	P4482-15DL	MYE4E3D	DUP	10/31/24 01:04	4X For Ba,Co,Ni,V	Jaswal	Confirms
232	P4482-14LDL	MYE4E3L	SD	10/31/24 01:07	20X For Ba,Co,Ni,V	Jaswal	Confirms
233	P4482-16DL	MYE4E3S	MS	10/31/24 01:10	4X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133210

Review By	Sarabjit Jaswal	Review On	10/30/2024 3:40:25 PM
Supervise By	Mohan Bera	Supervise On	10/30/2024 3:42:59 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

234	P4482-14MSDL	MYE4E3SRE	MS	10/31/24 01:14	4X For Ba,Co,Ni,V	Jaswal	Confirms
235	P4483-20DL	MYE4G9	SAM	10/31/24 01:17	4X For Ba,Co,Ni,V	Jaswal	Confirms
236	P4483-21DL	MYE4G9D	DUP	10/31/24 01:20	4X For Ba,Co,Ni,V	Jaswal	Confirms
237	P4483-20LDL	MYE4G9L	SD	10/31/24 01:24	20X For Ba,Co,Ni,V	Jaswal	Confirms
238	P4483-22DL	MYE4G9S	MS	10/31/24 01:27	4X For Ba,Co,Ni,V	Jaswal	Confirms
239	P4483-20MSDL	MYE4G9SRE	MS	10/31/24 01:30	4X For Ba,Co,Ni,V	Jaswal	Confirms
240	P4481-12A	MYE4C1ARE	PS	10/31/24 01:33	Not Required	Jaswal	Not Ok
241	P4482-14A	MYE4E3ARE	PS	10/31/24 01:37	PS For AG(0.02ML MP82908-10ML SAMPLE)	Jaswal	OK
242	P4483-20A	MYE4G9ARE	PS	10/31/24 01:40	PS For AG(0.02ML MP82908-10ML SAMPLE)	Jaswal	OK
243	CCV044	CCV044	CCV	10/31/24 01:49		Jaswal	OK
244	CCB044	CCB044	CCB	10/31/24 01:52		Jaswal	OK