

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE005	TUNE	11/10/24 22:06		Jaswal	OK
2	S00	S00	CAL1	11/10/24 23:04		Jaswal	OK
3	S02	S02	CAL3	11/10/24 23:11		Jaswal	OK
4	S03	S03	CAL4	11/10/24 23:14		Jaswal	OK
5	S04	S04	CAL5	11/10/24 23:17		Jaswal	OK
6	S05	S05	CAL6	11/10/24 23:20		Jaswal	OK
7	S06	S06	CAL7	11/10/24 23:23		Jaswal	OK
8	S07	S07	CAL8	11/10/24 23:25		Jaswal	OK
9	S08	S08	CAL9	11/10/24 23:28		Jaswal	OK
10	ICV005	ICV005	ICV	11/10/24 23:40	Inst. paused for calibration review	Jaswal	OK
11	ICB005	ICB005	ICB	11/10/24 23:44		Jaswal	OK
12	ICSA005	ICSA005	ICSA	11/10/24 23:47		Jaswal	OK
13	ICSAB005	ICSAB005	ICSAB	11/10/24 23:50		Jaswal	OK
14	CCV011	CCV011	CCV	11/11/24 00:07	Inst. paused for review	Jaswal	OK
15	CCB011	CCB011	CCB	11/11/24 00:10		Jaswal	OK
16	PB164599BL	PBS599	MB	11/11/24 00:14		Jaswal	OK
17	PB164599BS	LCS599	LCS	11/11/24 00:17		Jaswal	OK
18	PB164599BSD	LCSD599	LCSD	11/11/24 00:21		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

19	PB164601BL	PBS601	MB	11/11/24 00:24		Jaswal	OK
20	PB164601BS	LCS601	LCS	11/11/24 00:27		Jaswal	OK
21	PB164601BSD	LCSD601	LCSD	11/11/24 00:30		Jaswal	OK
22	PB164624BLRE	PBS624	SAM	11/11/24 00:34	Not Use	Jaswal	Not Ok
23	PB164624BSRE	LCS624	SAM	11/11/24 00:37	Not Use	Jaswal	Not Ok
24	PB164624BSDRE	LCSD624	SAM	11/11/24 00:40	Not Use	Jaswal	Not Ok
25	P4524-01	MYE4W0	SAM	11/11/24 00:43	INT_STD 45Sc(2),89Y(1) Fail	Jaswal	Dilution
26	P4524-02	MYE4W1	SAM	11/11/24 00:46	INT_STD 45Sc(2),89Y(1&2) Fail	Jaswal	Dilution
27	P4524-03	MYE4W2	SAM	11/11/24 00:49	INT_STD 45Sc(2),89Y(1) Fail	Jaswal	Dilution
28	P4524-04	MYE4W3	SAM	11/11/24 00:52	INT_STD 89Y(1) Fail	Jaswal	Dilution
29	P4524-05	MYE4W4	SAM	11/11/24 00:55	INT_STD 89Y(1) Fail	Jaswal	Dilution
30	P4524-06RE	MYE4W5	SAM	11/11/24 00:58	INT_STD 89Y(1) Fail,Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
31	P4524-07	MYE4W5D	DUP	11/11/24 01:01	INT_STD 89Y(1) Fail,Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
32	P4524-06L	MYE4W5L	SD	11/11/24 01:04	INT_STD 89Y(1) Fail,Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
33	P4524-08	MYE4W5S	MS	11/11/24 01:07	INT_STD 89Y(1) Fail,Ba,Co,Ni, V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

34	P4524-06MS	MYE4W5SRE	MS	11/11/24 01:10	MS Fail For As,Cu,Pb, INT_STD 89Y(1) Fail,Ba,Co,Ni, V High	Jaswal	Dilution
35	P4524-09	MYE4W6	SAM	11/11/24 01:13		Jaswal	OK
36	P4524-10	MYE4W7	SAM	11/11/24 01:16		Jaswal	OK
37	P4524-11	MYE4Y0	SAM	11/11/24 01:19		Jaswal	OK
38	P4524-12	MYE4Y1	SAM	11/11/24 01:22	INT_STD 89Y(1) Fail	Jaswal	Dilution
39	P4484-03A	MYE4H2ARE	PS	11/11/24 01:25	Not Use	Jaswal	Not Ok
40	P4484-03ADL2	MYE4H2ARE	PS	11/11/24 01:28	PS For As-0.0082 -M5801,Cr- 0.012 - M5658,Ni - 0.0075 M5961,V - 0.007 - M5697-PS For Ni,V,Cr	Jaswal	Confirms
41	P4524-13	MYE4Y2	SAM	11/11/24 01:31		Jaswal	OK
42	P4524-14	MYE4Y3	SAM	11/11/24 01:34		Jaswal	OK
43	P4524-15	MYE4Y4	SAM	11/11/24 01:37		Jaswal	OK
44	P4524-16	MYE4Y5	SAM	11/11/24 01:40		Jaswal	OK
45	P4524-17	MYE4Y6	SAM	11/11/24 01:43		Jaswal	OK
46	P4538-01	MYE4Y7	SAM	11/11/24 01:46		Jaswal	OK
47	CCV012	CCV012	CCV	11/11/24 01:49		Jaswal	OK
48	CCB012	CCB012	CCB	11/11/24 01:53		Jaswal	OK
49	P4538-02	MYE4Y8	SAM	11/11/24 01:57	INT_STD 89Y(1) Fail	Jaswal	Dilution
50	P4538-03	MYE4Y9	SAM	11/11/24 02:00		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

51	P4538-04	MYE4Z0	SAM	11/11/24 02:03		Jaswal	OK
52	P4538-05	MYE4Z1	SAM	11/11/24 02:06		Jaswal	OK
53	P4538-06	MYE4Z2	SAM	11/11/24 02:09		Jaswal	OK
54	P4538-07	MYE4Z3	SAM	11/11/24 02:12		Jaswal	OK
55	P4538-08	MYE4Z4	SAM	11/11/24 02:15		Jaswal	OK
56	P4538-09	MYE4Z5	SAM	11/11/24 02:18		Jaswal	OK
57	P4538-10	MYE4Z6	SAM	11/11/24 02:21	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
58	P4538-11	MYE4Z6D	DUP	11/11/24 02:24	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
59	P4538-10L	MYE4Z6L	SD	11/11/24 02:27	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
60	P4538-12	MYE4Z6S	MS	11/11/24 02:30	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
61	P4538-10MS	MYE4Z6SRE	MS	11/11/24 02:33	MS Fail For As,Pb,Ba,Co,Ni, V High	Jaswal	Dilution
62	P4538-13	MYE4Z7	SAM	11/11/24 02:36		Jaswal	OK
63	P4538-14	MYE4Z8	SAM	11/11/24 02:39		Jaswal	OK
64	P4538-15	MYE4Z9	SAM	11/11/24 02:42		Jaswal	OK
65	P4538-16	MYE500	SAM	11/11/24 02:45		Jaswal	OK
66	P4538-17	MYE501	SAM	11/11/24 02:48		Jaswal	OK
67	P4538-18	MYE502	SAM	11/11/24 02:51		Jaswal	OK
68	P4524-06	MYE4W5	SAM	11/11/24 02:53		Jaswal	OK
69	P4538-19	MYE503	SAM	11/11/24 02:56		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

70	P4538-20	MYE504	SAM	11/11/24 02:59		Jaswal	OK
71	P4538-21	MYE505	SAM	11/11/24 03:02		Jaswal	OK
72	P4541-01	MYE5B3	SAM	11/11/24 03:05		Jaswal	OK
73	P4541-02	MYE5B4	SAM	11/11/24 03:08		Jaswal	OK
74	P4541-03	MYE5B5	SAM	11/11/24 03:11		Jaswal	OK
75	CCV013	CCV013	CCV	11/11/24 03:15		Jaswal	OK
76	CCB013	CCB013	CCB	11/11/24 03:17		Jaswal	OK
77	P4484-03ADL	MYE4H2ARE	PS	11/11/24 03:24	PS For As-0.0082 -M5801,PS For As	Jaswal	Dilution
78	P4541-04	MYE5B6	SAM	11/11/24 03:27		Jaswal	OK
79	P4541-05	MYE5B7	SAM	11/11/24 03:30	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
80	P4541-06	MYE5B7D	DUP	11/11/24 03:33	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
81	P4541-05L	MYE5B7L	SD	11/11/24 03:36	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
82	P4541-07	MYE5B7S	MS	11/11/24 03:39	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
83	P4541-05MS	MYE5B7SRE	MS	11/11/24 03:42	MS Fail For Pb,Ba,Co,Ni, V High	Jaswal	Dilution
84	P4541-08	MYE5B8	SAM	11/11/24 03:45		Jaswal	OK
85	P4541-09	MYE5B9	SAM	11/11/24 03:48		Jaswal	OK
86	P4541-10	MYE5C0	SAM	11/11/24 03:51		Jaswal	OK
87	P4541-11	MYE5C1	SAM	11/11/24 03:54		Jaswal	OK
88	P4541-12	MYE5C2	SAM	11/11/24 03:57		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

89	P4541-13	MYE5C3	SAM	11/11/24 04:00		Jaswal	OK
90	P4541-14	MYE5C4	SAM	11/11/24 04:03		Jaswal	OK
91	P4541-15	MYE5C5	SAM	11/11/24 04:06		Jaswal	OK
92	P4541-16	MYE5C6	SAM	11/11/24 04:09		Jaswal	OK
93	P4541-17	MYE5C7	SAM	11/11/24 04:12		Jaswal	OK
94	P4541-18	MYE5C8	SAM	11/11/24 04:15		Jaswal	OK
95	P4541-19	MYE5C9	SAM	11/11/24 04:18		Jaswal	OK
96	P4541-20	MYE5D0	SAM	11/11/24 04:21		Jaswal	OK
97	P4541-21	MYE5D1	SAM	11/11/24 04:24		Jaswal	OK
98	P4541-22	MYE5D2	SAM	11/11/24 04:27		Jaswal	OK
99	P4521-08	MYE4P2	SAM	11/11/24 04:30	Not Use	Jaswal	Not Ok
100	P4521-09	MYE4P2D	DUP	11/11/24 04:33	Not Use	Jaswal	Not Ok
101	P4521-08L	MYE4P2L	SD	11/11/24 04:36	Not Use	Jaswal	Not Ok
102	P4521-10	MYE4P2S	MS	11/11/24 04:39	Not Use	Jaswal	Not Ok
103	P4521-08MS	MYE4P2SRE	MS	11/11/24 04:42	Not Use	Jaswal	Not Ok
104	CCV014	CCV014	CCV	11/11/24 04:45		Jaswal	OK
105	CCB014	CCB014	CCB	11/11/24 04:48		Jaswal	OK
106	P4521-11	MYE4P3	SAM	11/11/24 04:51	Not Use	Jaswal	Not Ok
107	P4521-12	MYE4P4	SAM	11/11/24 04:54	Not Use	Jaswal	Not Ok
108	P4521-13	MYE4P5	SAM	11/11/24 04:57	Not Use	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

109	P4521-14	MYE4Q2	SAM	11/11/24 05:00	Not Use	Jaswal	Not Ok
110	P4521-15	MYE4Q3	SAM	11/11/24 05:03	Not Use	Jaswal	Not Ok
111	P4521-16	MYE4Q4	SAM	11/11/24 05:06	Not Use	Jaswal	Not Ok
112	P4521-17	MYE4Q5	SAM	11/11/24 05:09	Not Use	Jaswal	Not Ok
113	P4521-18	MYE4Q6	SAM	11/11/24 05:12	Not Use	Jaswal	Not Ok
114	P4521-19	MYE4Q7	SAM	11/11/24 05:15	Not Use	Jaswal	Not Ok
115	P4521-20	MYE4Q8	SAM	11/11/24 05:18	Not Use	Jaswal	Not Ok
116	P4521-21	MYE4Q9	SAM	11/11/24 05:21	Not Use	Jaswal	Not Ok
117	P4521-22	MYE4R0	SAM	11/11/24 05:24	Not Use	Jaswal	Not Ok
118	P4524-06DL	MYE4W5	SAM	11/11/24 05:27	4X For Ba,Co,Ni,V	Jaswal	Confirms
119	P4524-07DL	MYE4W5D	DUP	11/11/24 05:30	4X For Ba,Co,Ni,V	Jaswal	Confirms
120	P4524-06LDL	MYE4W5L	SD	11/11/24 05:33	4X For Ba,Co,Ni,V	Jaswal	Confirms
121	P4524-08DL	MYE4W5S	MS	11/11/24 05:36	4X For Ba,Co,Ni,V	Jaswal	Confirms
122	P4524-06MSDL	MYE4W5SRE	MS	11/11/24 05:40	4X For Ba,Co,Ni,V	Jaswal	Confirms
123	P4538-10DL	MYE4Z6	SAM	11/11/24 05:42	4X For Ba,Co,Ni,V	Jaswal	Confirms
124	P4538-11DL	MYE4Z6D	DUP	11/11/24 05:46	4X For Ba,Co,Ni,V	Jaswal	Confirms
125	P4538-10LDL	MYE4Z6L	SD	11/11/24 05:49	4X For Ba,Co,Ni,V	Jaswal	Confirms
126	P4538-12DL	MYE4Z6S	MS	11/11/24 05:52	4X For Ba,Co,Ni,V	Jaswal	Confirms
127	P4538-10MSDL	MYE4Z6SRE	MS	11/11/24 05:55	4X For Ba,Co,Ni,V	Jaswal	Confirms
128	P4538-02DL	MYE4Y8	SAM	11/11/24 05:58	2X For INT_STD 89Y(1)	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

129	P4541-02DL	MYE5B4	SAM	11/11/24 06:01	Not Required	Jaswal	Not Ok
130	P4541-05DL	MYE5B7	SAM	11/11/24 06:04	4X For Ba,Co,Ni,V	Jaswal	Confirms
131	P4541-06DL	MYE5B7D	DUP	11/11/24 06:07	4X For Ba,Co,Ni,V	Jaswal	Confirms
132	P4541-05LDL	MYE5B7L	SD	11/11/24 06:10	4X For Ba,Co,Ni,V	Jaswal	Confirms
133	CCV015	CCV015	CCV	11/11/24 06:14		Jaswal	OK
134	CCB015	CCB015	CCB	11/11/24 06:16		Jaswal	OK
135	P4541-07DL	MYE5B7S	MS	11/11/24 06:20	4X For Ba,Co,Ni,V	Jaswal	Confirms
136	P4541-05MSDL	MYE5B7SRE	MS	11/11/24 06:23	4X For Ba,Co,Ni,V	Jaswal	Confirms
137	P4521-08DL	MYE4P2	SAM	11/11/24 06:26	4X For Ba,Co,Ni,V	Jaswal	Confirms
138	P4521-09DL	MYE4P2D	DUP	11/11/24 06:29	4X For Ba,Co,Ni,V	Jaswal	Confirms
139	P4521-08LDL	MYE4P2L	SD	11/11/24 06:32	4X For Ba,Co,Ni,V	Jaswal	Confirms
140	P4521-10DL	MYE4P2S	MS	11/11/24 06:35	4X For Ba,Co,Ni,V	Jaswal	Confirms
141	P4521-08MSDL	MYE4P2SRE	MS	11/11/24 06:38	4X For Ba,Co,Ni,V	Jaswal	Confirms
142	PB164624BL	PBS624	MB	11/11/24 06:41		Jaswal	OK
143	PB164624BS	LCS624	LCS	11/11/24 06:44		Jaswal	OK
144	PB164624BSD	LCSD624	LCSD	11/11/24 06:48		Jaswal	OK
145	PB164625BL	PBS625	MB	11/11/24 06:51		Jaswal	OK
146	PB164625BS	LCS625	LCS	11/11/24 06:54		Jaswal	OK
147	PB164625BSD	LCSD625	LCSD	11/11/24 06:57		Jaswal	OK
148	PB164798BL	PBW798	MB	11/11/24 07:01		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

149	PB164798BS	LCS798	LCS	11/11/24 07:04		Jaswal	OK
150	P4542-04	MYE5D6	SAM	11/11/24 07:07	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
151	P4542-05	MYE5D6D	DUP	11/11/24 07:10	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
152	P4542-04L	MYE5D6L	SD	11/11/24 07:13	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
153	P4542-06	MYE5D6S	MS	11/11/24 07:16	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
154	P4542-04MS	MYE5D6SRE	MS	11/11/24 07:19	MS Fail For - As,Pb,Ba,Co,Ni, V High	Jaswal	Dilution
155	P4542-04DL	MYE5D6	SAM	11/11/24 07:22	4X For Ba,Co,Ni,V	Jaswal	Confirms
156	P4542-05DL	MYE5D6D	DUP	11/11/24 07:25	4X For Ba,Co,Ni,V	Jaswal	Confirms
157	P4542-04LDL	MYE5D6L	SD	11/11/24 07:28	4X For Ba,Co,Ni,V	Jaswal	Confirms
158	P4542-06DL	MYE5D6S	MS	11/11/24 07:32	4X For Ba,Co,Ni,V	Jaswal	Confirms
159	P4542-04MSDL	MYE5D6SRE	MS	11/11/24 07:35	4X For Ba,Co,Ni,V	Jaswal	Confirms
160	P4542-01	MYE5D3	SAM	11/11/24 07:38		Jaswal	OK
161	CCV016	CCV016	CCV	11/11/24 07:41		Jaswal	OK
162	CCB016	CCB016	CCB	11/11/24 07:44		Jaswal	OK
163	P4542-02	MYE5D4	SAM	11/11/24 07:47		Jaswal	OK
164	P4542-03	MYE5D5	SAM	11/11/24 07:50		Jaswal	OK
165	P4542-07	MYE5D7	SAM	11/11/24 07:53		Jaswal	OK
166	P4542-08	MYE5D8	SAM	11/11/24 07:56		Jaswal	OK
167	P4542-09	MYE5D9	SAM	11/11/24 07:59		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

168	P4542-10	MYE5E0	SAM	11/11/24 08:02		Jaswal	OK
169	P4542-11	MYE5E1	SAM	11/11/24 08:05		Jaswal	OK
170	P4542-12	MYE5E2	SAM	11/11/24 08:08		Jaswal	OK
171	P4542-13	MYE5E3	SAM	11/11/24 08:11		Jaswal	OK
172	P4542-14	MYE5E4	SAM	11/11/24 08:14		Jaswal	OK
173	P4542-15	MYE5E5	SAM	11/11/24 08:17		Jaswal	OK
174	P4542-16	MYE5E6	SAM	11/11/24 08:20		Jaswal	OK
175	P4542-17	MYE5E7	SAM	11/11/24 08:23		Jaswal	OK
176	P4542-18	MYE5E8	SAM	11/11/24 08:26		Jaswal	OK
177	P4542-19	MYE5E9	SAM	11/11/24 08:29		Jaswal	OK
178	P4542-20	MYE5F0	SAM	11/11/24 08:32		Jaswal	OK
179	P4542-21	MYE5F1	SAM	11/11/24 08:35		Jaswal	OK
180	P4542-22	MYE5F2	SAM	11/11/24 08:38		Jaswal	OK
181	PB164766BL	PBW766	MB	11/11/24 08:41		Jaswal	OK
182	PB164766BS	LCS766	LCS	11/11/24 08:45		Jaswal	OK
183	P4730-01	MC0B82	SAM	11/11/24 08:48	Cr High	Jaswal	Dilution
184	P4730-02	MC0B83	SAM	11/11/24 08:51		Jaswal	OK
185	P4730-03	MC0B84	SAM	11/11/24 08:54		Jaswal	OK
186	P4730-04	MC0B86	SAM	11/11/24 08:58		Jaswal	OK
187	P4730-05	MC0B86D	DUP	11/11/24 09:01		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

188	P4730-04L	MC0B86L	SD	11/11/24 09:04		Jaswal	OK
189	ICSA006	ICSA006	ICSA	11/11/24 09:07		Jaswal	OK
190	ICSAB006	ICSAB006	ICSAB	11/11/24 09:10		Jaswal	OK
191	CCV017	CCV017	CCV	11/11/24 09:13		Jaswal	OK
192	CCB017	CCB017	CCB	11/11/24 09:16		Jaswal	OK
193	P4730-06	MC0B86S	MS	11/11/24 09:19		Jaswal	OK
194	P4730-07	MC0B88	SAM	11/11/24 09:22		Jaswal	OK
195	P4580-01	MYE5L3	SAM	11/11/24 09:25		Jaswal	OK
196	P4580-02	MYE5L4	SAM	11/11/24 09:29		Jaswal	OK
197	P4580-03	MYE5L5	SAM	11/11/24 09:32		Jaswal	OK
198	P4580-04	MYE5L6	SAM	11/11/24 09:35		Jaswal	OK
199	P4580-05	MYE5L7	SAM	11/11/24 09:38		Jaswal	OK
200	P4580-06	MYE5L8	SAM	11/11/24 09:42		Jaswal	OK
201	P4580-07	MYE5L9	SAM	11/11/24 09:45		Jaswal	OK
202	P4580-08	MYE5M0	SAM	11/11/24 09:48		Jaswal	OK
203	P4580-09	MYE5M1	SAM	11/11/24 09:51		Jaswal	OK
204	P4580-10	MYE5M2	SAM	11/11/24 09:55		Jaswal	OK
205	P4542-14DL	MYE5E4	SAM	11/11/24 09:58	NOT USED	Jaswal	Not Ok
206	PB164627BL	PBS627	MB	11/11/24 10:01		Jaswal	OK
207	PB164627BS	LCS627	LCS	11/11/24 10:04		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

208	PB164627BSD	LCSD627	LCSD	11/11/24 10:08		Jaswal	OK
209	P4543-01	MYE5F3	SAM	11/11/24 10:11		Jaswal	OK
210	P4543-02	MYE5F4	SAM	11/11/24 10:14		Jaswal	OK
211	P4543-03	MYE5F5	SAM	11/11/24 10:17	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
212	P4543-04	MYE5F5D	DUP	11/11/24 10:20	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
213	P4543-03L	MYE5F5L	SD	11/11/24 10:23	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
214	P4543-05	MYE5F5S	MS	11/11/24 10:26	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
215	P4543-03MS	MYE5F5SRE	MS	11/11/24 10:29	MS Fail For - As(109.84ppb),Pb(83.23ppb),Ag(Below RL),Ba,Co,Ni, V High	Jaswal	Dilution
216	P4543-06	MYE5F6	SAM	11/11/24 10:32		Jaswal	OK
217	P4543-07	MYE5F7	SAM	11/11/24 10:35		Jaswal	OK
218	P4543-08	MYE5F8	SAM	11/11/24 10:38		Jaswal	OK
219	CCV018	CCV018	CCV	11/11/24 10:41		Jaswal	OK
220	CCB018	CCB018	CCB	11/11/24 10:44		Jaswal	OK
221	P4543-09	MYE5F9	SAM	11/11/24 10:47		Jaswal	OK
222	P4543-10	MYE5G0	SAM	11/11/24 10:50		Jaswal	OK
223	P4543-11	MYE5G2	SAM	11/11/24 10:53		Jaswal	OK
224	P4543-12	MYE5G3	SAM	11/11/24 10:56		Jaswal	OK
225	P4543-13	MYE506	SAM	11/11/24 10:59		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

226	P4543-14	MYE507	SAM	11/11/24 11:02		Jaswal	OK
227	P4543-15	MYE508	SAM	11/11/24 11:05		Jaswal	OK
228	P4543-16	MYE509	SAM	11/11/24 11:08		Jaswal	OK
229	P4543-17	MYE510	SAM	11/11/24 11:11		Jaswal	OK
230	P4543-18	MYE511	SAM	11/11/24 11:14		Jaswal	OK
231	P4543-19	MYE512	SAM	11/11/24 11:17		Jaswal	OK
232	P4543-20	MYE513	SAM	11/11/24 11:20		Jaswal	OK
233	P4543-21	MYE514	SAM	11/11/24 11:23		Jaswal	OK
234	P4543-22	MYE515	SAM	11/11/24 11:26		Jaswal	OK
235	P4543-03DL	MYE5F5	SAM	11/11/24 11:29	4X For Ba,Co,Ni,V	Jaswal	Confirms
236	P4543-04DL	MYE5F5D	DUP	11/11/24 11:33	4X For Ba,Co,Ni,V	Jaswal	Confirms
237	P4543-03LDL	MYE5F5L	SD	11/11/24 11:36	4X For Ba,Co,Ni,V	Jaswal	Confirms
238	P4543-05DL	MYE5F5S	MS	11/11/24 11:39	4X For Ba,Co,Ni,V	Jaswal	Confirms
239	P4543-03MSDL	MYE5F5SRE	MS	11/11/24 11:42	4X For Ba,Co,Ni,V	Jaswal	Confirms
240	P4524-06A	MYE4W5ARE	PS	11/11/24 11:45	PS For As - 0.0037 - M5801 , Cu - 0.011 - M5697, Pb - 0.0033 M6026	Jaswal	OK
241	P4538-10A	MYE4Z6ARE	PS	11/11/24 11:48	PS For As - 0.0041 - M5801, Pb - 0.029 MP83156	Jaswal	OK
242	P4541-05A	MYE5B7ARE	PS	11/11/24 11:51	PS For Pb - 0.037 - MP83156	Jaswal	OK
243	P4542-04A	MYE5D6ARE	PS	11/11/24 11:54	PS For As - 0.02 MP83228, Pb - 0.023 MP83156	Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

244	PB164632BL	PBS632	MB	11/11/24 11:57		Jaswal	OK
245	PB164632BS	LCS632	LCS	11/11/24 12:00		Jaswal	OK
246	CCV019	CCV019	CCV	11/11/24 12:04	Ca Fail	Jaswal	OK
247	CCB019	CCB019	CCB	11/11/24 12:06		Jaswal	OK
248	PB164632BSD	LCSD632	LCSD	11/11/24 12:10		Jaswal	OK
249	P4570-04	MYE537	SAM	11/11/24 12:22		Jaswal	OK
250	P4570-05	MYE538	SAM	11/11/24 12:25		Jaswal	OK
251	P4570-06	MYE539	SAM	11/11/24 12:28		Jaswal	OK
252	P4570-07	MYE540	SAM	11/11/24 12:31		Jaswal	OK
253	P4570-08	MYE541	SAM	11/11/24 12:34		Jaswal	OK
254	P4570-09	MYE542	SAM	11/11/24 12:37		Jaswal	OK
255	P4570-10	MYE563	SAM	11/11/24 12:40		Jaswal	OK
256	P4570-11	MYE564	SAM	11/11/24 12:43		Jaswal	OK
257	P4570-12	MYE565	SAM	11/11/24 12:47		Jaswal	OK
258	P4570-13	MYE566	SAM	11/11/24 12:50		Jaswal	OK
259	P4570-14	MYE567	SAM	11/11/24 12:53		Jaswal	OK
260	P4570-15	MYE568	SAM	11/11/24 12:56		Jaswal	OK
261	P4570-16	MYE569	SAM	11/11/24 12:59		Jaswal	OK
262	P4570-17	MYE570	SAM	11/11/24 13:02		Jaswal	OK
263	P4570-18	MYE571	SAM	11/11/24 13:05		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

264	P4570-19	MYE572	SAM	11/11/24 13:08		Jaswal	OK
265	P4570-20	MYE573	SAM	11/11/24 13:11		Jaswal	OK
266	P4570-21	MYE574	SAM	11/11/24 13:14		Jaswal	OK
267	P4570-22	MYE575	SAM	11/11/24 13:17		Jaswal	OK
268	P4570-01	MYE531	SAM	11/11/24 13:20	Ba,Co,Ni, V High in MS RE	Jaswal	Dilution
269	P4570-02	MYE531D	DUP	11/11/24 13:23	Ba,Co,Ni, V High in MS RE	Jaswal	Dilution
270	P4570-01L	MYE531L	SD	11/11/24 13:26	Ba,Co,Ni, V High in MS RE	Jaswal	Dilution
271	P4570-03	MYE531S	MS	11/11/24 13:29	Ba,Co,Ni, V High in MS RE	Jaswal	Dilution
272	CCV020	CCV020	CCV	11/11/24 13:32		Jaswal	OK
273	CCB020	CCB020	CCB	11/11/24 13:35		Jaswal	OK
274	P4570-01MS	MYE531SRE	MS	11/11/24 13:38	MS Fail For Ag(Below RL),Ba,Co,Ni, V High	Jaswal	Dilution
275	P4543-21DL	MYE514	SAM	11/11/24 13:41		Jaswal	Confirms
276	P4570-01DL	MYE531	SAM	11/11/24 13:55	4X For Ba,Co,Ni,V	Jaswal	Confirms
277	P4570-02DL	MYE531D	DUP	11/11/24 13:58	4X For Ba,Co,Ni,V	Jaswal	Confirms
278	P4570-031LDL	MYE531S	SD	11/11/24 14:01	20X For Ba,Co,Ni,V	Jaswal	Confirms
279	P4570-03DL	MYE531S	MS	11/11/24 14:04	4X For Ba,Co,Ni,V	Jaswal	Confirms
280	P4570-01MSDL	MYE531SRE	MS	11/11/24 14:07	4X For Ba,Co,Ni,V	Jaswal	Confirms
281	CCV021	CCV021	CCV	11/11/24 14:10		Jaswal	OK
282	CCB021	CCB021	CCB	11/11/24 14:13		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

283	PB164630BL	PBS630	MB	11/11/24 14:19		Jaswal	OK
284	PB164630BS	LCS630	LCS	11/11/24 14:23		Jaswal	OK
285	PB164630BSD	LCSD630	LCSD	11/11/24 14:26		Jaswal	OK
286	PB164635BL	PBS635	MB	11/11/24 14:32		Jaswal	OK
287	PB164635BS	LCS635	LCS	11/11/24 14:35		Jaswal	OK
288	PB164635BSD	LCSD635	LCSD	11/11/24 14:39		Jaswal	OK
289	P4569-01	MYE516	SAM	11/11/24 14:42		Jaswal	OK
290	P4569-02	MYE517	SAM	11/11/24 14:45		Jaswal	OK
291	P4569-03	MYE518	SAM	11/11/24 14:48		Jaswal	OK
292	P4569-04	MYE519	SAM	11/11/24 14:51	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
293	P4569-05	MYE519D	DUP	11/11/24 14:54	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
294	P4569-04L	MYE519L	SD	11/11/24 14:57	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
295	P4569-06	MYE519S	MS	11/11/24 15:00	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
296	P4569-04MS	MYE519SRE	MS	11/11/24 15:03	MS Fail For Ag(1.05ppb),Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
297	P4569-07	MYE520	SAM	11/11/24 15:06		Jaswal	OK
298	P4569-08	MYE521	SAM	11/11/24 15:09		Jaswal	OK
299	P4569-09	MYE522	SAM	11/11/24 15:12		Jaswal	OK
300	P4569-10	MYE523	SAM	11/11/24 15:15		Jaswal	OK
301	P4569-11	MYE524	SAM	11/11/24 15:18		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

302	P4569-12	MYE525	SAM	11/11/24 15:21		Jaswal	OK
303	P4569-13	MYE526	SAM	11/11/24 15:24		Jaswal	OK
304	P4569-14	MYE527	SAM	11/11/24 15:27		Jaswal	OK
305	P4569-15	MYE528	SAM	11/11/24 15:30		Jaswal	OK
306	P4569-16	MYE529	SAM	11/11/24 15:33		Jaswal	OK
307	P4569-17	MYE530	SAM	11/11/24 15:36		Jaswal	OK
308	CCV022	CCV022	CCV	11/11/24 15:42		Jaswal	OK
309	CCB022	CCB022	CCB	11/11/24 15:45		Jaswal	OK
310	P4569-18	MYE532	SAM	11/11/24 15:48		Jaswal	OK
311	P4569-19	MYE533	SAM	11/11/24 15:51		Jaswal	OK
312	P4569-20	MYE534	SAM	11/11/24 15:54		Jaswal	OK
313	P4569-21	MYE535	SAM	11/11/24 15:57		Jaswal	OK
314	P4569-22	MYE536	SAM	11/11/24 16:00		Jaswal	OK
315	P4571-01	MYE543	SAM	11/11/24 16:03	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
316	P4571-02	MYE543D	DUP	11/11/24 16:06	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
317	P4571-01L	MYE543L	SD	11/11/24 16:09	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
318	P4571-03	MYE543S	MS	11/11/24 16:13	Ba,Co,Ni, V High in MSRE	Jaswal	Dilution
319	P4571-01MS	MYE543SRE	MS	11/11/24 16:16	MS Fail For Pb(51.79ppb),Ba,Co,Ni, V High	Jaswal	Dilution
320	P4571-04	MYE576	SAM	11/11/24 16:18		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133382

Review By	Sarabjit Jaswal	Review On	11/12/2024 5:31:27 PM
Supervise By	Mohan Bera	Supervise On	11/13/2024 2:51:14 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

321	P4571-07	MYE579	SAM	11/11/24 16:21		Jaswal	OK
322	P4571-08	MYE580	SAM	11/11/24 16:24		Jaswal	OK
323	P4571-09	MYE581	SAM	11/11/24 16:27		Jaswal	OK
324	P4571-10	MYE544	SAM	11/11/24 16:30		Jaswal	OK
325	P4571-11	MYE545	SAM	11/11/24 16:33		Jaswal	OK
326	P4571-12	MYE546	SAM	11/11/24 16:36		Jaswal	OK
327	P4571-13	MYE547	SAM	11/11/24 16:40		Jaswal	OK
328	P4571-14	MYE548	SAM	11/11/24 16:42		Jaswal	OK
329	P4571-15	MYE549	SAM	11/11/24 16:46		Jaswal	OK
330	P4571-16	MYE550	SAM	11/11/24 16:49		Jaswal	OK
331	P4571-17	MYE551	SAM	11/11/24 16:52		Jaswal	OK
332	P4571-18	MYE552	SAM	11/11/24 16:55		Jaswal	OK
333	P4571-19	MYE553	SAM	11/11/24 16:58		Jaswal	OK
334	P4571-20	MYE554	SAM	11/11/24 17:01	Ni High	Jaswal	Dilution
335	P4571-21	MYE555	SAM	11/11/24 17:04		Jaswal	OK
336	P4571-22	MYE556	SAM	11/11/24 17:07		Jaswal	OK
337	CCV023	CCV023	CCV	11/11/24 17:10		Jaswal	OK
338	CCB023	CCB023	CCB	11/11/24 17:12		Jaswal	OK