



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P4490

Project: Pre Treatment Plant

RunNo.: LB133071

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.500	0.50	100	90-110	10/23/2024
Sample ID: CCV1 Phosphorus, Total	mg/L	0.501	0.50	100	90-110	10/23/2024
Sample ID: CCV2 Phosphorus, Total	mg/L	0.501	0.50	100	90-110	10/23/2024

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P4490

Project: Pre Treatment Plant

RunNo.: LB133073

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.500	0.50	100	90-110	10/23/2024
Sample ID: CCV1 Orthophosphate as P	mg/L	0.502	0.5	100	90-110	10/23/2024
Sample ID: CCV2 Orthophosphate as P	mg/L	0.502	0.5	100	90-110	10/23/2024

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P4490

Project: Pre Treatment Plant

RunNo.: LB133173

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1.1	1	110	90-110	10/28/2024
Sample ID: CCV1 Ammonia as N	mg/L	0.99	1	99	90-110	10/28/2024
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	10/28/2024
Sample ID: CCV3 Ammonia as N	mg/L	1.1	1	110	90-110	10/28/2024
Sample ID: CCV4 Ammonia as N	mg/L	1.1	1	110	90-110	10/28/2024