

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4490

Project: Pre Treatment Plant

Run No.: LB133064

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133064BS							
TSS	mg/L	550	532		97	1	90-110	10/22/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4490

Project: Pre Treatment Plant

Run No.: LB133073

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133073BS							
Orthophosphate as P	mg/L	0.5	0.50		100	1	90-110	10/23/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4490

Project: Pre Treatment Plant

Run No.: LB133080

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133080BS							
BOD5	mg/L	198	192		97	1	84.6-115.4	10/23/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4490

Project: Pre Treatment Plant

Run No.: LB133159

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133159BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	10/28/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4490

Project: Pre Treatment Plant

Run No.: LB133071

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB164344BS							
Phosphorus, Total	mg/L	0.50	0.50		100	1	90-110	10/23/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4490

Project: Pre Treatment Plant

Run No.: LB133173

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB164383BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	10/28/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P4490

Project: Pre Treatment Plant

Run No.: LB133173

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB164463BS							
Ammonia as N	mg/Kg	50	50.4		101	1	90-110	10/28/2024