



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **Aramark Uniforms**

**Project Name: Monthly 2024**

**Project # N/A**

**Chemtech Project # P4491**

**Test Name: Metals ICP-Group1,Mercury**

### **A. Number of Samples and Date of Receipt:**

2 Water samples were received on 10/23/2024.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: BOD5, Mercury, Metals Group1, Metals ICP-Group1, TPH and TSS. This data package contains results for Metals ICP-Group1,Mercury.

### **C. Analytical Techniques:**

The analysis of Metals ICP-Group1 was based on method 6010D, digestion based on method 3010 (waters). The analysis and digestion of Mercury was based on method 7470A.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (34839-40MS) analysis met criteria for all samples except for Mercury due to sample matrix interference. The Matrix Spike (MW-1MS) analysis met criteria for all samples except for Zinc due to Chemical Interference during Digestion Process.

The Matrix Spike Duplicate (34839-40MSD) analysis met criteria for all samples except for Mercury due to sample matrix interference. The Matrix Spike Duplicate (MW-1MSD) analysis met criteria for all samples except for Zinc due to Chemical Interference during Digestion Process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

### **E. Additional Comments:**

As per special requirement for this project form-1 and Hit Summary are reported in mg/l.

---



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_