


LB133129

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	0.007	0.007	-26	0.000	-26	0				0.40
0.2	0.200	0.224	0.024	709	0.0 %	709					0.2
2.5	2.500	2.301	-0.199	7747	0.0 %	7747					0.8
5.0	5.000	5.261	0.261	17775	0.0 %	17775					5
7.5	7.500	7.498	-0.002	25354	0.0 %	25354					0
10.0	10.000	9.909	-0.091	33522	0.0 %	33522					1

LB133129 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Oct 2024 11:32:12

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-26	-	0.0000	7470A	PPB	25 Oct 2024 11:51:23	S	Std
0.2 - 1	50.2	709	-	0.2000	7470A	PPB	25 Oct 2024 11:53:40	S	Std
2.5 - 1	52.5	7747	-	2.5000	7470A	PPB	25 Oct 2024 11:55:57	S	Std
5.0 - 1	55	17775	-	5.0000	7470A	PPB	25 Oct 2024 11:58:14	S	Std
7.5 - 1	57.5	25354	-	7.5000	7470A	PPB	25 Oct 2024 12:03:01	S	Std
10.0 - 1	57.0	33522	-	10.0000	7470A	PPB	25 Oct 2024 12:07:56	S	Std
ICV53 - 1	ICV53	14370	4.2560	-7470A	PPB	25 Oct 2024 12:21:25	U	SMPL	
ICB53 - 1	ICB53	-137	-0.0256	-7470A	PPB	25 Oct 2024 12:23:41	U	SMPL	
CCV73 - 1	CCV73	17307	5.1229	-7470A	PPB	25 Oct 2024 12:25:57	U	SMPL	
CCB73 - 1	CCB73	-77	-0.0079	-7470A	PPB	25 Oct 2024 12:30:50	U	SMPL	
CRA - 1	CRA	651	0.2069	-7470A	PPB	25 Oct 2024 12:33:05	U	SMPL	
HighStd - 1	HighStd	32392	9.5751	-7470A	PPB	25 Oct 2024 12:35:20	U	SMPL	
ChkStd - 1	ChkStd	22804	6.7453	-7470A	PPB	25 Oct 2024 12:37:35	U	SMPL	
PB164407BL - 1	PBW	-95	-0.0132	-7470A	PPB	25 Oct 2024 12:42:31	U	SMPL	
PB164407BS - 1	LCSW	13635	4.0391	-7470A	PPB	25 Oct 2024 12:44:46	U	SMPL	
P3390-07 - 1	LOD-MDL-WATER-01-QT3-2024	153	0.0600	-7470A	PPB	25 Oct 2024 12:49:37	U	SMPL	
P3390-08 - 1	LOQ-WATER-02-QT3-2024	592	0.1895	-7470A	PPB	25 Oct 2024 12:51:53	U	SMPL	
P4460-06 - 1	WB-303-SW	162	0.0626	-7470A	PPB	25 Oct 2024 12:54:10	U	SMPL	
P4485-03 - 1	D20241001-02-03	3485	1.0434	-7470A	PPB	25 Oct 2024 12:56:27	U	SMPL	
P4491-02 - 1	COMP	277	0.0966	-7470A	PPB	25 Oct 2024 12:58:42	U	SMPL	
CCV74 - 1	CCV74	16807	4.9753	-7470A	PPB	25 Oct 2024 13:00:59	U	SMPL	
CCB74 - 1	CCB74	-40	0.0030	-7470A	PPB	25 Oct 2024 13:06:17	U	SMPL	
P4546-03 - 1	34839-40	881	0.2748	-7470A	PPB	25 Oct 2024 13:14:14	U	SMPL	
P4546-03DUP - 1	34839-40DUP	831	0.2601	-7470A	PPB	25 Oct 2024 13:19:03	U	SMPL	
P4546-03MS - 1	34839-40MS	8505	2.5250	-7470A	PPB	25 Oct 2024 13:21:18	U	SMPL	
P4546-03MSD - 1	34839-40MSD	7630	2.2668	-7470A	PPB	25 Oct 2024 13:23:34	U	SMPL	
P4546-03LX5 - 1		-4	0.0136	-7470A	PPB	25 Oct 2024 13:25:51	U	SMPL	
P4546-03A - 1		14802	4.3835	-7470A	PPB	25 Oct 2024 13:28:09	U	SMPL	
CCV75 - 1	CCV75	17437	5.1612	-7470A	PPB	25 Oct 2024 13:30:25	U	SMPL	
CCB75 - 1	CCB75	-151	-0.0298	-7470A	PPB	25 Oct 2024 13:32:43	U	SMPL	