

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** P4491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4491 **SAS No.:** P4491
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV53	Mercury	4.26	4.0	106	90 - 110	CV	10/25/2024	12:21	LB133129

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** P4491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4491 **SAS No.:** P4491
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV73	Mercury	5.12	5.0	102	90 - 110	CV	10/25/2024	12:25	LB133129
CCV74	Mercury	4.98	5.0	100	90 - 110	CV	10/25/2024	13:00	LB133129
CCV75	Mercury	5.16	5.0	103	90 - 110	CV	10/25/2024	13:30	LB133129

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** P4491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4491 **SAS No.:** P4491
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	478	510	94	90 - 110	P	10/29/2024	11:59	LB133198
	Copper	525	510	103	90 - 110	P	10/29/2024	11:59	LB133198
	Lead	958	1000	96	90 - 110	P	10/29/2024	11:59	LB133198
	Nickel	514	530	97	90 - 110	P	10/29/2024	11:59	LB133198
	Zinc	1040	1000	104	90 - 110	P	10/29/2024	11:59	LB133198

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** P4491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4491 **SAS No.:** P4491
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.68	6.0	95	80 - 120	P	10/29/2024	12:31	LB133198
	Copper	21.7	20.0	108	80 - 120	P	10/29/2024	12:31	LB133198
	Lead	9.73	12.0	81	80 - 120	P	10/29/2024	12:31	LB133198
	Nickel	40.0	40.0	100	80 - 120	P	10/29/2024	12:31	LB133198
	Zinc	38.9	40.0	97	80 - 120	P	10/29/2024	12:31	LB133198

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** P4491

Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4491 **SAS No.:** P4491

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2450	2500	98	90 - 110	P	10/29/2024	13:01	LB133198
	Copper	1250	1250	100	90 - 110	P	10/29/2024	13:01	LB133198
	Lead	4870	5000	97	90 - 110	P	10/29/2024	13:01	LB133198
	Nickel	2450	2500	98	90 - 110	P	10/29/2024	13:01	LB133198
	Zinc	2510	2500	100	90 - 110	P	10/29/2024	13:01	LB133198
CCV02	Cadmium	2330	2500	93	90 - 110	P	10/29/2024	14:04	LB133198
	Copper	1200	1250	96	90 - 110	P	10/29/2024	14:04	LB133198
	Lead	4670	5000	93	90 - 110	P	10/29/2024	14:04	LB133198
	Nickel	2340	2500	94	90 - 110	P	10/29/2024	14:04	LB133198
	Zinc	2390	2500	96	90 - 110	P	10/29/2024	14:04	LB133198
CCV03	Cadmium	2660	2500	106	90 - 110	P	10/29/2024	15:02	LB133198
	Copper	1130	1250	90	90 - 110	P	10/29/2024	15:02	LB133198
	Lead	4700	5000	94	90 - 110	P	10/29/2024	15:02	LB133198
	Nickel	2580	2500	103	90 - 110	P	10/29/2024	15:02	LB133198
	Zinc	2660	2500	106	90 - 110	P	10/29/2024	15:02	LB133198
CCV04	Cadmium	2640	2500	106	90 - 110	P	10/29/2024	15:59	LB133198
	Copper	1230	1250	98	90 - 110	P	10/29/2024	15:59	LB133198
	Lead	4620	5000	92	90 - 110	P	10/29/2024	15:59	LB133198
	Nickel	2540	2500	102	90 - 110	P	10/29/2024	15:59	LB133198
	Zinc	2610	2500	104	90 - 110	P	10/29/2024	15:59	LB133198
CCV05	Cadmium	2630	2500	105	90 - 110	P	10/29/2024	16:53	LB133198
	Copper	1230	1250	98	90 - 110	P	10/29/2024	16:53	LB133198
	Lead	4590	5000	92	90 - 110	P	10/29/2024	16:53	LB133198
	Nickel	2530	2500	101	90 - 110	P	10/29/2024	16:53	LB133198
	Zinc	2280	2500	91	90 - 110	P	10/29/2024	16:53	LB133198
CCV06	Cadmium	2600	2500	104	90 - 110	P	10/29/2024	17:53	LB133198
	Copper	1200	1250	96	90 - 110	P	10/29/2024	17:53	LB133198
	Lead	4570	5000	91	90 - 110	P	10/29/2024	17:53	LB133198
	Nickel	2490	2500	100	90 - 110	P	10/29/2024	17:53	LB133198
	Zinc	2670	2500	107	90 - 110	P	10/29/2024	17:53	LB133198
CCV07	Cadmium	2320	2500	93	90 - 110	P	10/29/2024	18:20	LB133198
	Copper	1270	1250	102	90 - 110	P	10/29/2024	18:20	LB133198
	Lead	4600	5000	92	90 - 110	P	10/29/2024	18:20	LB133198
	Nickel	2290	2500	92	90 - 110	P	10/29/2024	18:20	LB133198

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** P4491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4491 **SAS No.:** P4491
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Zinc	2260	2500	90	90 - 110	P	10/29/2024	18:20	LB133198
