

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Aramark Uniforms **SDG No.:** P4491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4491 **SAS No.:** P4491
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	6.66	1.0	666	-5	7	10/29/2024	12:44	LB133198
	Copper	-5.53	2.0	276	-18	22	10/29/2024	12:44	LB133198
	Lead	7.65			-12	12	10/29/2024	12:44	LB133198
	Nickel	2.53	2.0	126	-38	42	10/29/2024	12:44	LB133198
	Zinc	-0.39			-40	40	10/29/2024	12:44	LB133198
ICSAB01	Cadmium	970	972	100	826	1120	10/29/2024	12:48	LB133198
	Copper	496	511	97	434	588	10/29/2024	12:48	LB133198
	Lead	54.7	49.0	112	37	61	10/29/2024	12:48	LB133198
	Nickel	1020	954	107	810	1100	10/29/2024	12:48	LB133198
	Zinc	1090	952	114	809	1095	10/29/2024	12:48	LB133198