

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>P4491</u>
contract: <u>ARAM01</u>	lab code: <u>CHEM</u>	case no.: <u>P4491</u> sas no.: <u>P4491</u>
matrix: <u>Water</u>	sample id: <u>P4546-03</u>	client id: <u>34839-40MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>P4546-03MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	82 - 119	2.53		0.28		4.0	56	N	CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Aramark Uniforms **level:** low **sdg no.:** P4491
contract: ARAM01 **lab code:** CHEM **case no.:** P4491 **sas no.:** P4491
matrix: Water **sample id:** P4546-03 **client id:** 34839-40MSD
Percent Solids for Sample: NA **Spiked ID:** P4546-03MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	82 - 119	2.27		0.28		4.0	50	N	CV

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Aramark Uniforms</u>	level: <u>low</u>	sdg no.: <u>P4491</u>
contract: <u>ARAM01</u>	lab code: <u>CHEM</u>	case no.: <u>P4491</u> sas no.: <u>P4491</u>
matrix: <u>Water</u>	sample id: <u>P4548-01</u>	client id: <u>MW-1MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>P4548-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	92.7		2.09	J	100	91		P
Copper	ug/L	75 - 125	138		10.0	U	150	92		P
Lead	ug/L	75 - 125	388		6.00	U	500	78		P
Nickel	ug/L	75 - 125	243		6.80	J	250	95		P
Zinc	ug/L	75 - 125	298		254		100	43	N	P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Aramark Uniforms **level:** low **sdg no.:** P4491
contract: ARAM01 **lab code:** CHEM **case no.:** P4491 **sas no.:** P4491
matrix: Water **sample id:** P4548-01 **client id:** MW-1MSD
Percent Solids for Sample: NA **Spiked ID:** P4548-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	95.0		2.09	J	100	93		P
Copper	ug/L	75 - 125	141		10.0	U	150	94		P
Lead	ug/L	75 - 125	398		6.00	U	500	80		P
Nickel	ug/L	75 - 125	249		6.80	J	250	97		P
Zinc	ug/L	75 - 125	303		254		100	49	N	P