

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: P4491

Contract: ARAM01

Lab Code: CHEM

Case No.: P4491

SAS No.: P4491

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164407BS Mercury	ug/L	4.0	4.04		101	82 - 119	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: P4491

Contract: ARAM01

Lab Code: CHEM

Case No.: P4491

SAS No.: P4491

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164438BS							
Cadmium	ug/L	100	84.8		85	80 - 120	P
Copper	ug/L	150	155		103	80 - 120	P
Lead	ug/L	500	426		85	80 - 120	P
Nickel	ug/L	250	225		90	80 - 120	P
Zinc	ug/L	100	89.1		89	80 - 120	P