

## **CASE NARRATIVE**

### **Chemtech Consulting Group**

**Project Name: NJ Soil PT**

**Project # N/A**

**Chemtech Project # P4495**

**Test Name: PESTICIDE Group3**

### **A. Number of Samples and Date of Receipt:**

25 Solid samples were received on 10/23/2024.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group1, Anions Group2, Corrosivity, Cyanide, Diesel Range Organics, EPH, Flash Point, Gasoline Range Organics, Herbicide Group1, Hexavalent Chromium, Mercury, Metals Group3, Metals ICP-Group1, Nitrite, Oil and Grease, PCB, PESTICIDE Group1, PESTICIDE Group2, PESTICIDE Group3, Phosphorus, Total, SVOCMS Group1, SVOCMS Group2, SVOCMS Group3, SVOCMS Group4, TKN, TOC, TS and VOCMS Group1. This data package contains results for PESTICIDE Group3.

### **C. Analytical Techniques:**

The analysis was performed on instrument ECD\_L. The front column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11 The rear column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0. 5 um df,: Catalog # 7HM-G016-17. .The analysis of PESTICIDE Group3s was based on method 8081B and extraction was done based on method 3541.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

Sample PT-TXP-SOIL was diluted due to high concentration.

### **E. Additional Comments:**

### **F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_