

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

Continuing Calib Date: 11/14/2024 Initial Calibration Date(s): 11/13/2024 11/13/2024

Continuing Calib Time: 20:40 Initial Calibration Time(s): 11:44 13:23

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.43	7.43	7.33	7.53	0.00
MCPD	7.61	7.61	7.51	7.71	0.00
2,4-DCAA	7.24	7.24	7.14	7.34	0.00
Dalapon	2.63	2.64	2.54	2.74	0.01
MCPA	7.76	7.76	7.66	7.86	0.00
DICHLORPROP	8.14	8.14	8.04	8.24	0.00
2,4-D	8.37	8.37	8.27	8.47	0.00
2,4,5-TP(Silvex)	9.25	9.25	9.15	9.35	0.00
2,4,5-T	9.54	9.54	9.44	9.64	0.00
2,4-DB	10.12	10.12	10.02	10.22	0.00
Dinoseb	11.33	11.33	11.23	11.43	0.00
Pentachlorophenol	8.67	8.67	8.57	8.77	0.00
4-Nitrophenol	7.04	7.04	6.94	7.14	0.00
PICLORAM	11.14	11.14	11.04	11.24	0.00
DCPA	11.62	11.63	11.53	11.73	0.01
3,5-DICHLOROBENZ	6.41	6.41	6.31	6.51	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

Continuing Calib Date: 11/14/2024 Initial Calibration Date(s): 11/13/2024 11/13/2024

Continuing Calib Time: 20:40 Initial Calibration Time(s): 11:44 13:23

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.94	7.94	7.84	8.04	0.00
MCPD	8.04	8.05	7.95	8.15	0.01
2,4-DCAA	7.74	7.74	7.64	7.84	0.00
Dalapon	2.69	2.70	2.60	2.80	0.01
MCPA	8.29	8.29	8.19	8.39	0.00
DICHLORPROP	8.66	8.66	8.56	8.76	0.01
2,4-D	8.99	8.99	8.89	9.09	0.00
2,4,5-TP(Silvex)	9.89	9.90	9.80	10.00	0.01
2,4,5-T	10.31	10.32	10.22	10.42	0.01
2,4-DB	10.88	10.88	10.78	10.98	0.00
Dinoseb	11.26	11.26	11.16	11.36	0.00
Pentachlorophenol	9.52	9.52	9.42	9.62	0.00
4-Nitrophenol	7.27	7.27	7.17	7.37	0.01
PICLORAM	12.36	12.36	12.26	12.46	0.00
DCPA	12.31	12.31	12.21	12.41	0.00
3,5-DICHLOROBENZ	6.69	6.70	6.60	6.80	0.01