

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

Continuing Calib Date: 11/25/2024 Initial Calibration Date(s): 11/13/2024 11/13/2024

Continuing Calib Time: 10:14 Initial Calibration Time(s): 11:44 13:23

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.41	7.43	7.33	7.53	0.02
MCP	7.60	7.61	7.51	7.71	0.01
2,4-DCAA	7.23	7.24	7.14	7.34	0.01
Dalapon	2.64	2.64	2.54	2.74	0.01
MCPA	7.75	7.76	7.66	7.86	0.01
DICHLORPROP	8.12	8.14	8.04	8.24	0.02
2,4-D	8.35	8.37	8.27	8.47	0.02
2,4,5-TP(Silvex)	9.23	9.25	9.15	9.35	0.02
2,4,5-T	9.52	9.54	9.44	9.64	0.02
2,4-DB	10.10	10.12	10.02	10.22	0.02
Dinoseb	11.31	11.33	11.23	11.43	0.02
Pentachlorophenol	8.65	8.67	8.57	8.77	0.02
4-Nitrophenol	7.03	7.04	6.94	7.14	0.01
PICLORAM	11.11	11.14	11.04	11.24	0.03
DCPA	11.60	11.63	11.53	11.73	0.03
3,5-DICHLOROBENZ	6.40	6.41	6.31	6.51	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

Continuing Calib Date: 11/25/2024 Initial Calibration Date(s): 11/13/2024 11/13/2024

Continuing Calib Time: 10:14 Initial Calibration Time(s): 11:44 13:23

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.93	7.94	7.84	8.04	0.01
MCP	8.03	8.05	7.95	8.15	0.02
2,4-DCAA	7.73	7.74	7.64	7.84	0.01
Dalapon	2.69	2.70	2.60	2.80	0.01
MCPA	8.27	8.29	8.19	8.39	0.02
DICHLORPROP	8.64	8.66	8.56	8.76	0.02
2,4-D	8.97	8.99	8.89	9.09	0.02
2,4,5-TP(Silvex)	9.88	9.90	9.80	10.00	0.02
2,4,5-T	10.30	10.32	10.22	10.42	0.02
2,4-DB	10.86	10.88	10.78	10.98	0.02
Dinoseb	11.24	11.26	11.16	11.36	0.02
Pentachlorophenol	9.50	9.52	9.42	9.62	0.02
4-Nitrophenol	7.25	7.27	7.17	7.37	0.02
PICLORAM	12.34	12.36	12.26	12.46	0.02
DCPA	12.28	12.31	12.21	12.41	0.03
3,5-DICHLOROBENZ	6.68	6.70	6.60	6.80	0.02