

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL01 Date Analyzed: 11/14/2024

Lab Sample No.: HSTDCCC750 Data File : PS028534.D Time Analyzed: 12:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.543	9.443	9.643	755.870	712.500	6.1
2,4,5-TP(Silvex)	9.249	9.150	9.350	759.220	712.500	6.6
2,4-D	8.369	8.270	8.470	736.520	705.000	4.5
2,4-DB	10.118	10.019	10.219	773.480	712.500	8.6
2,4-DCAA	7.240	7.141	7.341	762.420	750.000	1.7
3,5-DICHLOROBENZOIC ACID	6.410	6.311	6.511	713.710	697.500	2.3
4-Nitrophenol	7.039	6.939	7.139	678.040	682.500	-0.7
Dalapon	2.634	2.539	2.739	656.070	682.500	-3.9
DCPA	11.625	11.526	11.726	700.890	720.000	-2.7
DICAMBA	7.428	7.329	7.529	751.910	705.000	6.7
DICHLORPROP	8.138	8.039	8.239	707.800	705.000	0.4
Dinoseb	11.332	11.233	11.433	756.040	705.000	7.2
MCPA	7.762	7.663	7.863	74.130	69.750	6.3
MCPP	7.612	7.512	7.712	80.190	70.500	13.7
Pentachlorophenol	8.669	8.570	8.770	793.700	712.500	11.4
PICLORAM	11.138	11.039	11.239	722.400	712.500	1.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL01 Date Analyzed: 11/14/2024

Lab Sample No.: HSTDCCC750 Data File : PS028534.D Time Analyzed: 12:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.314	10.216	10.416	759.140	712.500	6.5
2,4,5-TP(Silvex)	9.893	9.796	9.996	766.920	712.500	7.6
2,4-D	8.987	8.890	9.090	744.150	705.000	5.6
2,4-DB	10.882	10.784	10.984	752.580	712.500	5.6
2,4-DCAA	7.740	7.642	7.842	765.530	750.000	2.1
3,5-DICHLOROBENZOIC ACID	6.693	6.595	6.795	713.760	697.500	2.3
4-Nitrophenol	7.266	7.168	7.368	675.750	682.500	-1.0
Dalapon	2.693	2.599	2.799	630.950	682.500	-7.6
DCPA	12.306	12.208	12.408	694.870	720.000	-3.5
DICAMBA	7.940	7.842	8.042	758.700	705.000	7.6
DICHLORPROP	8.657	8.559	8.759	722.790	705.000	2.5
Dinoseb	11.261	11.163	11.363	752.810	705.000	6.8
MCPA	8.288	8.191	8.391	71.570	69.750	2.6
MCPP	8.043	7.945	8.145	76.500	70.500	8.5
Pentachlorophenol	9.516	9.419	9.619	777.580	712.500	9.1
PICLORAM	12.357	12.260	12.460	730.530	712.500	2.5