

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL02 Date Analyzed: 11/14/2024

Lab Sample No.: HSTDCCC750 Data File : PS028545.D Time Analyzed: 20:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.541	9.443	9.643	770.840	712.500	8.2
2,4,5-TP(Silvex)	9.248	9.150	9.350	775.590	712.500	8.9
2,4-D	8.368	8.270	8.470	754.130	705.000	7.0
2,4-DB	10.117	10.019	10.219	792.690	712.500	11.3
2,4-DCAA	7.239	7.141	7.341	779.080	750.000	3.9
3,5-DICHLOROBENZOIC ACID	6.410	6.311	6.511	728.310	697.500	4.4
4-Nitrophenol	7.038	6.939	7.139	690.420	682.500	1.2
Dalapon	2.633	2.539	2.739	671.270	682.500	-1.6
DCPA	11.624	11.526	11.726	718.820	720.000	-0.2
DICAMBA	7.427	7.329	7.529	767.520	705.000	8.9
DICHLORPROP	8.137	8.039	8.239	724.010	705.000	2.7
Dinoseb	11.330	11.233	11.433	777.230	705.000	10.2
MCPA	7.761	7.663	7.863	75.310	69.750	8.0
MCPP	7.611	7.512	7.712	80.840	70.500	14.7
Pentachlorophenol	8.668	8.570	8.770	809.590	712.500	13.6
PICLORAM	11.137	11.039	11.239	729.270	712.500	2.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL02 Date Analyzed: 11/14/2024

Lab Sample No.: HSTDCCC750 Data File : PS028545.D Time Analyzed: 20:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.313	10.216	10.416	771.980	712.500	8.3
2,4,5-TP(Silvex)	9.892	9.796	9.996	780.280	712.500	9.5
2,4-D	8.986	8.890	9.090	757.930	705.000	7.5
2,4-DB	10.880	10.784	10.984	763.510	712.500	7.2
2,4-DCAA	7.739	7.642	7.842	775.870	750.000	3.4
3,5-DICHLOROBENZOIC ACID	6.692	6.595	6.795	723.600	697.500	3.7
4-Nitrophenol	7.265	7.168	7.368	688.130	682.500	0.8
Dalapon	2.694	2.599	2.799	645.900	682.500	-5.4
DCPA	12.305	12.208	12.408	702.620	720.000	-2.4
DICAMBA	7.939	7.842	8.042	772.000	705.000	9.5
DICHLORPROP	8.655	8.559	8.759	727.180	705.000	3.1
Dinoseb	11.260	11.163	11.363	770.740	705.000	9.3
MCPA	8.287	8.191	8.391	72.210	69.750	3.5
MCPP	8.042	7.945	8.145	77.010	70.500	9.2
Pentachlorophenol	9.516	9.419	9.619	791.850	712.500	11.1
PICLORAM	12.356	12.260	12.460	735.290	712.500	3.2