

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL03 Date Analyzed: 11/25/2024

Lab Sample No.: HSTDCCC750 Data File : PS028614.D Time Analyzed: 10:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.521	9.443	9.643	764.730	712.500	7.3
2,4,5-TP(Silvex)	9.229	9.150	9.350	772.700	712.500	8.4
2,4-D	8.352	8.270	8.470	746.500	705.000	5.9
2,4-DB	10.095	10.019	10.219	722.880	712.500	1.5
2,4-DCAA	7.226	7.141	7.341	786.480	750.000	4.9
3,5-DICHLOROBENZOIC ACID	6.398	6.311	6.511	747.950	697.500	7.2
4-Nitrophenol	7.025	6.939	7.139	731.160	682.500	7.1
Dalapon	2.635	2.539	2.739	722.540	682.500	5.9
DCPA	11.599	11.526	11.726	773.680	720.000	7.5
DICAMBA	7.414	7.329	7.529	775.860	705.000	10.1
DICHLORPROP	8.121	8.039	8.239	754.340	705.000	7.0
Dinoseb	11.305	11.233	11.433	751.080	705.000	6.5
MCPA	7.746	7.663	7.863	74.370	69.750	6.6
MCPP	7.596	7.512	7.712	75.810	70.500	7.5
Pentachlorophenol	8.650	8.570	8.770	813.670	712.500	14.2
PICLORAM	11.113	11.039	11.239	641.340	712.500	-10.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL03 Date Analyzed: 11/25/2024

Lab Sample No.: HSTDCCC750 Data File : PS028614.D Time Analyzed: 10:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.295	10.216	10.416	712.820	712.500	0.0
2,4,5-TP(Silvex)	9.876	9.796	9.996	738.720	712.500	3.7
2,4-D	8.971	8.890	9.090	693.470	705.000	-1.6
2,4-DB	10.862	10.784	10.984	640.620	712.500	-10.1
2,4-DCAA	7.726	7.642	7.842	712.370	750.000	-5.0
3,5-DICHLOROBENZOIC ACID	6.683	6.595	6.795	679.230	697.500	-2.6
4-Nitrophenol	7.254	7.168	7.368	673.810	682.500	-1.3
Dalapon	2.694	2.599	2.799	638.130	682.500	-6.5
DCPA	12.284	12.208	12.408	759.250	720.000	5.5
DICAMBA	7.926	7.842	8.042	721.200	705.000	2.3
DICHLORPROP	8.642	8.559	8.759	687.480	705.000	-2.5
Dinoseb	11.241	11.163	11.363	693.240	705.000	-1.7
MCPA	8.274	8.191	8.391	62.820	69.750	-9.9
MCPP	8.029	7.945	8.145	64.370	70.500	-8.7
Pentachlorophenol	9.499	9.419	9.619	759.150	712.500	6.5
PICLORAM	12.335	12.260	12.460	612.260	712.500	-14.1