

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL04 Date Analyzed: 11/25/2024

Lab Sample No.: HSTDCCC750 Data File : PS028618.D Time Analyzed: 11:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.520	9.443	9.643	772.000	712.500	8.4
2,4,5-TP(Silvex)	9.228	9.150	9.350	772.960	712.500	8.5
2,4-D	8.351	8.270	8.470	748.480	705.000	6.2
2,4-DB	10.095	10.019	10.219	739.400	712.500	3.8
2,4-DCAA	7.226	7.141	7.341	760.440	750.000	1.4
3,5-DICHLOROBENZOIC ACID	6.400	6.311	6.511	747.220	697.500	7.1
4-Nitrophenol	7.025	6.939	7.139	731.460	682.500	7.2
Dalapon	2.635	2.539	2.739	716.670	682.500	5.0
DCPA	11.598	11.526	11.726	773.240	720.000	7.4
DICAMBA	7.414	7.329	7.529	772.410	705.000	9.6
DICHLORPROP	8.121	8.039	8.239	752.250	705.000	6.7
Dinoseb	11.305	11.233	11.433	746.680	705.000	5.9
MCPA	7.747	7.663	7.863	72.770	69.750	4.3
MCPP	7.597	7.512	7.712	76.060	70.500	7.9
Pentachlorophenol	8.650	8.570	8.770	813.880	712.500	14.2
PICLORAM	11.113	11.039	11.239	708.400	712.500	-0.6

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/13/2024 11/13/2024

Client Sample No.: CCAL04 Date Analyzed: 11/25/2024

Lab Sample No.: HSTDCCC750 Data File : PS028618.D Time Analyzed: 11:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.296	10.216	10.416	730.860	712.500	2.6
2,4,5-TP(Silvex)	9.877	9.796	9.996	749.410	712.500	5.2
2,4-D	8.972	8.890	9.090	704.430	705.000	-0.1
2,4-DB	10.863	10.784	10.984	664.870	712.500	-6.7
2,4-DCAA	7.727	7.642	7.842	727.560	750.000	-3.0
3,5-DICHLOROBENZOIC ACID	6.684	6.595	6.795	690.120	697.500	-1.1
4-Nitrophenol	7.254	7.168	7.368	694.190	682.500	1.7
Dalapon	2.695	2.599	2.799	670.080	682.500	-1.8
DCPA	12.285	12.208	12.408	773.750	720.000	7.5
DICAMBA	7.926	7.842	8.042	732.130	705.000	3.8
DICHLORPROP	8.642	8.559	8.759	693.330	705.000	-1.7
Dinoseb	11.242	11.163	11.363	697.790	705.000	-1.0
MCPA	8.274	8.191	8.391	62.280	69.750	-10.7
MCPP	8.030	7.945	8.145	63.310	70.500	-10.2
Pentachlorophenol	9.500	9.419	9.619	772.470	712.500	8.4
PICLORAM	12.335	12.260	12.460	694.580	712.500	-2.5