

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS111324

Review By	yogesh	Review On	11/14/2024 8:46:40 AM
Supervise By	Ankita	Supervise On	11/14/2024 11:10:35 AM
SubDirectory	PS111324	HP Acquire Method	HP Processing Method PS111324

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP23458,PP23459,PP23460,PP23461,PP23462
CCC	PP23462
Internal Standard/PEM	
ICV/I.BLK	PP23469
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028484.D	13 Nov 2024 08:41		AR\AJ	Ok
2	I.BLK	I.BLK	PS028485.D	13 Nov 2024 09:06	Need ICAL	AR\AJ	Not Ok
3	HSTDCCC750	HSTDCCC750	PS028486.D	13 Nov 2024 10:56	Need ICAL	AR\AJ	Not Ok
4	I.BLK	I.BLK	PS028487.D	13 Nov 2024 11:20		AR\AJ	Ok
5	HSTDICC200	HSTDICC200	PS028488.D	13 Nov 2024 11:44		AR\AJ	Ok
6	HSTDICC500	HSTDICC500	PS028489.D	13 Nov 2024 12:09		AR\AJ	Ok
7	HSTDICC750	HSTDICC750	PS028490.D	13 Nov 2024 12:34		AR\AJ	Ok
8	HSTDICC1000	HSTDICC1000	PS028491.D	13 Nov 2024 12:58		AR\AJ	Ok
9	HSTDICC1500	HSTDICC1500	PS028492.D	13 Nov 2024 13:23		AR\AJ	Ok
10	HSTDICV750	ICVPS111324	PS028493.D	13 Nov 2024 13:48		AR\AJ	Ok
11	I.BLK	I.BLK	PS028494.D	13 Nov 2024 14:12		AR\AJ	Ok
12	HSTDCCC750	HSTDCCC750	PS028495.D	13 Nov 2024 14:37		AR\AJ	Ok
13	P4793-01	M00-24-00345	PS028496.D	13 Nov 2024 15:02		AR\AJ	Ok,M
14	P4795-01	LAW-23-00189	PS028497.D	13 Nov 2024 15:26		AR\AJ	Ok,M
15	PB164863BL	PB164863BL	PS028498.D	13 Nov 2024 15:51		AR\AJ	Ok
16	PB164863BS	PB164863BS	PS028499.D	13 Nov 2024 16:15		AR\AJ	Ok
17	P4756-01	BP-B4	PS028500.D	13 Nov 2024 16:40		AR\AJ	Ok,M
18	P4756-01MS	BP-B4MS	PS028501.D	13 Nov 2024 17:05	Some compound recovery fail	AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS111324

Review By	yogesh	Review On	11/14/2024 8:46:40 AM
Supervise By	Ankita	Supervise On	11/14/2024 11:10:35 AM
SubDirectory	PS111324	HP Acquire Method	HP Processing Method PS111324
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP23458,PP23459,PP23460,PP23461,PP23462		
CCC	PP23462		
Internal Standard/PEM			
ICV/I.BLK	PP23469		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P4756-01MSD	BP-B4MSD	PS028502.D	13 Nov 2024 17:29	some compound recovery fail	AR\AJ	Ok,M
20	P4798-01	MH-6	PS028503.D	13 Nov 2024 17:54		AR\AJ	Ok,M
21	P4807-01	TP-7	PS028504.D	13 Nov 2024 18:18		AR\AJ	Ok
22	P4807-05	BF-F14	PS028505.D	13 Nov 2024 18:42		AR\AJ	Ok
23	I.BLK	I.BLK	PS028506.D	13 Nov 2024 19:07		AR\AJ	Ok
24	HSTDCCC750	HSTDCCC750	PS028507.D	13 Nov 2024 19:32		AR\AJ	Ok
25	P4807-09	BF-F13	PS028508.D	13 Nov 2024 19:56		AR\AJ	Ok
26	P4807-13	TP-6	PS028509.D	13 Nov 2024 20:21		AR\AJ	Ok
27	P4809-01	MH-5	PS028510.D	13 Nov 2024 20:45		AR\AJ	Ok,M
28	PB164946BL	PB164946BL	PS028511.D	13 Nov 2024 21:10		AR\AJ	Ok
29	PB164946BS	PB164946BS	PS028512.D	13 Nov 2024 21:35		AR\AJ	Ok,M
30	P4788-01MS	BP-G3MS	PS028513.D	13 Nov 2024 21:59	Some compound recovery fail	AR\AJ	Ok,M
31	P4788-01MSD	BP-G3MSD	PS028514.D	13 Nov 2024 22:24	Some compound recovery fail	AR\AJ	Ok,M
32	I.BLK	I.BLK	PS028515.D	13 Nov 2024 22:48		AR\AJ	Ok
33	HSTDCCC750	HSTDCCC750	PS028516.D	13 Nov 2024 23:13		AR\AJ	Ok
34	PB164959BL	PB164959BL	PS028517.D	13 Nov 2024 23:37		AR\AJ	Ok
35	PB164959BS	PB164959BS	PS028518.D	14 Nov 2024 00:02	Recovery Fail in DINOSEB-I	AR\AJ	Ok
36	PB164880TB	PB164880TB	PS028519.D	14 Nov 2024 00:26		AR\AJ	Ok
37	P4799-03	WC-TA2-02-C	PS028520.D	14 Nov 2024 00:51		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS111324

Review By	yogesh	Review On	11/14/2024 8:46:40 AM
Supervise By	Ankita	Supervise On	11/14/2024 11:10:35 AM
SubDirectory	PS111324	HP Acquire Method	HP Processing Method PS111324
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP23458,PP23459,PP23460,PP23461,PP23462		
CCC Internal Standard/PEM	PP23462		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469		

38	P4799-03MS	WC-TA2-02-CMS	PS028521.D	14 Nov 2024 01:15	Some compound recovery fail	AR\AJ	Ok,M
39	P4799-03MSD	WC-TA2-02-CMSD	PS028522.D	14 Nov 2024 01:40	Some compound recovery fail	AR\AJ	Ok,M
40	P4799-07	WC-TA2-03-C	PS028523.D	14 Nov 2024 02:05		AR\AJ	Ok
41	P4799-11	WC-TA1-01-C	PS028524.D	14 Nov 2024 02:29	will be reanalyzed for confirmation	AR\AJ	Not Ok
42	P4799-15	WC-TA1-02-C	PS028525.D	14 Nov 2024 02:54		AR\AJ	Ok
43	P4799-19	WC-TA1-03-C	PS028526.D	14 Nov 2024 03:18		AR\AJ	Ok
44	I.BLK	I.BLK	PS028527.D	14 Nov 2024 03:43		AR\AJ	Ok
45	HSTDCCC750	HSTDCCC750	PS028528.D	14 Nov 2024 04:08		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS111424

Review By	yogesh	Review On	11/15/2024 10:07:20 AM
Supervise By	Ankita	Supervise On	11/15/2024 10:08:05 AM
SubDirectory	PS111424	HP Acquire Method	HP Processing Method PS111324

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP23458,PP23459,PP23460,PP23461,PP23462
CCC Internal Standard/PEM	PP23462
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028529.D	14 Nov 2024 08:57		AR\AJ	Ok
2	I.BLK	I.BLK	PS028530.D	14 Nov 2024 09:21		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS028531.D	14 Nov 2024 09:45		AR\AJ	Ok
4	P4799-11	WC-TA1-01-C	PS028532.D	14 Nov 2024 11:09		AR\AJ	Ok
5	I.BLK	I.BLK	PS028533.D	14 Nov 2024 11:38		AR\AJ	Ok
6	HSTDCCC750	HSTDCCC750	PS028534.D	14 Nov 2024 12:02		AR\AJ	Ok
7	PB164971BL	PB164971BL	PS028535.D	14 Nov 2024 16:35		AR\AJ	Ok
8	PB164971BS	PB164971BS	PS028536.D	14 Nov 2024 17:00		AR\AJ	Ok
9	P4821-01	BP-F-24	PS028537.D	14 Nov 2024 17:24		AR\AJ	Ok
10	P4821-05	BP-F-2	PS028538.D	14 Nov 2024 17:49		AR\AJ	Ok
11	P4821-05MS	BP-F-2MS	PS028539.D	14 Nov 2024 18:14	Some compound recovery fail	AR\AJ	Ok,M
12	P4821-05MSD	BP-F-2MSD	PS028540.D	14 Nov 2024 18:38	Some compound recovery fail	AR\AJ	Ok,M
13	P4823-01	MH-4	PS028541.D	14 Nov 2024 19:02		AR\AJ	Ok,M
14	P4833-01	MH-731	PS028542.D	14 Nov 2024 19:27		AR\AJ	Ok
15	P4833-05	TP-2	PS028543.D	14 Nov 2024 19:51		AR\AJ	Ok
16	I.BLK	I.BLK	PS028544.D	14 Nov 2024 20:16		AR\AJ	Ok
17	HSTDCCC750	HSTDCCC750	PS028545.D	14 Nov 2024 20:40		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS112524

Review By	yogesh	Review On	11/26/2024 8:21:25 AM
Supervise By	Ankita	Supervise On	11/26/2024 2:25:39 PM
SubDirectory	PS112524	HP Acquire Method	HP Processing Method PS111324
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP23458,PP23459,PP23460,PP23461,PP23462		
CCC	PP23462		
Internal Standard/PEM			
ICV/I.BLK	PP23469		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS028612.D	25 Nov 2024 09:26		AR\AJ	Ok
2	I.BLK	I.BLK	PS028613.D	25 Nov 2024 09:50		AR\AJ	Ok,M
3	HSTDCCC750	HSTDCCC750	PS028614.D	25 Nov 2024 10:14		AR\AJ	Ok,M
4	P4495-17	PT-HERB-SOIL	PS028615.D	25 Nov 2024 10:39	Surrogate Fail in 2,4-DCAA-I	AR\AJ	ReRun
5	P4495-17RE	PT-HERB-SOILRE	PS028616.D	25 Nov 2024 11:03	Surrogate Fail in 2,4-DCAA-I	AR\AJ	Confirms
6	I.BLK	I.BLK	PS028617.D	25 Nov 2024 11:27		AR\AJ	Ok,M
7	HSTDCCC750	HSTDCCC750	PS028618.D	25 Nov 2024 11:51		AR\AJ	Ok,M
8	PB165228BL	PB165228BL	PS028619.D	25 Nov 2024 16:41		AR\AJ	Ok,M
9	PB165228BS	PB165228BS	PS028620.D	25 Nov 2024 17:05		AR\AJ	Ok,M
10	P4938-01	MH-732	PS028621.D	25 Nov 2024 17:29		AR\AJ	Ok,M
11	P4938-01MS	MH-732MS	PS028622.D	25 Nov 2024 17:53	Some compound recovery fail	AR\AJ	Ok,M
12	P4938-01MSD	MH-732MSD	PS028623.D	25 Nov 2024 18:17	Some compound recovery fail	AR\AJ	Ok,M
13	P4938-05	MH-734	PS028624.D	25 Nov 2024 18:41		AR\AJ	Ok,M
14	P4951-01	AU-05-112124	PS028625.D	25 Nov 2024 19:05	Surrogate Fail in 2,4-DCAA-II	AR\AJ	ReRun
15	P4985-01	MH-756-WC	PS028626.D	25 Nov 2024 19:29		AR\AJ	Ok,M
16	P4985-05	MH-740-WC	PS028627.D	25 Nov 2024 19:53	Surrogate Fail in 2,4-DCAA-I	AR\AJ	ReRun
17	I.BLK	I.BLK	PS028628.D	25 Nov 2024 20:17		AR\AJ	Ok,M
18	HSTDCCC750	HSTDCCC750	PS028629.D	25 Nov 2024 21:30		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS112524

Review By	yogesh	Review On	11/26/2024 8:21:25 AM
Supervise By	Ankita	Supervise On	11/26/2024 2:25:39 PM
SubDirectory	PS112524	HP Acquire Method	HP Processing Method PS111324
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP23458,PP23459,PP23460,PP23461,PP23462		
CCC Internal Standard/PEM	PP23462		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469		

M : Manual Integration