

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4495

Client: Chemtech Consulting Group

Analytical Method: 8151A Datafile : PS028536.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB164971BS	DICAMBA	166.6	169	ug/Kg	101				72	129	
	MCPD	16.66	16.2	ug/Kg	97				70	130	
	Dalapon	166.6	151	ug/Kg	91				70	130	
	MCPA	16.66	15.9	ug/Kg	95				70	130	
	DICHLORPROP	166.6	170	ug/Kg	102				77	135	
	2,4-D	166.6	169	ug/Kg	101				65	144	
	2,4,5-TP(Silvex)	166.6	173	ug/Kg	104				74	146	
	2,4,5-T	166.6	172	ug/Kg	103				77	134	
	2,4-DB	166.6	166	ug/Kg	100				72	122	
	Dinoseb	166.6	163	ug/Kg	98				74	132	
	Pentachlorophenol	166.6	182	ug/Kg	109				70	130	
	4-Nitrophenol	166.6	160	ug/Kg	96				70	130	
	PICLORAM	166.6	141	ug/Kg	85				70	130	
	DCPA	166.6	174	ug/Kg	104				70	130	
	3,5-DICHLOROBENZOIC	166.6	169	ug/Kg	101				70	130	