

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM **Case No.:** P4495 **SAS No.:** P4495 **SDG NO.:** P4495

Instrument ID: ECD_O **Calibration Date(s):** 10/15/2024 10/16/2024
Calibration Times: 18:27 02:36

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO107184.D</u>		CF 750 = <u>PO107185.D</u>			
CF 500 = <u>PO107186.D</u>		CF 250 = <u>PO107187.D</u>		CF 050 = <u>PO107188.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	252081604	261702867	271939076	291118488	272229420	269814291 5
Aroclor-1016-2	(2)	373760062	383028583	397887352	419075308	410673320	396884925 5
Aroclor-1016-3	(3)	234678633	244145249	256575754	276405664	244042480	251169556 6
Aroclor-1016-4	(4)	185622493	193848624	202963902	215081552	165543020	192611918 10
Aroclor-1016-5	(5)	176326470	182141932	191487930	203514892	159796500	182653545 9
Aroclor-1260-1	(1)	238480871	246497651	259155974	277220676	271587560	258588546 6
Aroclor-1260-2	(2)	243402558	251747185	264083512	284216460	272650860	263220115 6
Aroclor-1260-3	(3)	167353086	171560129	181491762	195138068	183182660	179745141 6
Aroclor-1260-4	(4)	162125557	166999933	175289716	187582032	185522120	175503872 6
Aroclor-1260-5	(5)	268686175	273623828	283035894	299271412	297351500	284393762 5
Decachlorobiphenyl		2362750210	2427138920	2496479000	2603827440	2376686800	2453376474 4
Tetrachloro-m-xylene		8902656430	9087809293	9285762860	9548006280	8745262400	9113899453 3
Aroclor-1242-1	(1)	207947063	205572916	222208808	238956456	243522880	223641625 8
Aroclor-1242-2	(2)	302439626	299949616	323550412	342804408	348858880	323520588 7
Aroclor-1242-3	(3)	191912697	193178427	209638152	224306728	225847640	208976729 8
Aroclor-1242-4	(4)	150447174	145338196	162797508	173135964	168308740	160005516 7
Aroclor-1242-5	(5)	141009343	143408436	153689098	164623016	179679780	156481935 10
Decachlorobiphenyl		2339420350	2427742867	2492700840	2551827400	2426410200	2447620331 3
Tetrachloro-m-xylene		8860043620	8651743680	9146784160	9504075440	9174828800	9067495140 4
Aroclor-1248-1	(1)	155913014	164800668	175191332	184609772	188867720	173876501 8
Aroclor-1248-2	(2)	222845576	238202488	253637968	270074056	280366620	253025342 9
Aroclor-1248-3	(3)	231904923	246233752	259706092	274388212	266197800	255686156 7
Aroclor-1248-4	(4)	237210289	246633865	260533272	272872732	271860120	257822056 6
Aroclor-1248-5	(5)	236273499	246591072	260733536	276314120	280304600	260043365 7
Decachlorobiphenyl		2339780580	2371533013	2504089680	2566158680	2396701200	2435652631 4
Tetrachloro-m-xylene		8719403610	9119992693	9358904180	9492784200	8995773400	9137371617 3
Aroclor-1254-1	(1)	246822697	256579683	264992242	286189960	289157060	268748328 7
Aroclor-1254-2	(2)	348120894	361641555	373031238	401237980	402995760	377405485 6
Aroclor-1254-3	(3)	336293322	345653409	355909306	378498460	372630440	357796987 5
Aroclor-1254-4	(4)	213138598	220171503	227187688	242530220	244046960	229414994 6
Aroclor-1254-5	(5)	194609638	201069208	205869028	220139672	216546640	207646837 5
Decachlorobiphenyl		2357746210	2401897693	2508594000	2601578200	2359367400	2445836701 4
Tetrachloro-m-xylene		8883763230	9164627320	9184369060	9508828480	8791509400	9106619498 3
Aroclor-1268-1	(1)	351862425	352413819	364622844	381760528	369380480	364008019 3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	313096403	314491445	324862806	339637696	320299560	322477582	3
Aroclor-1268-3	(3)	272236453	272760276	282620036	293538700	275661940	279363481	3
Aroclor-1268-4	(4)	112236308	111261111	114714594	116428084	101458340	111219687	5
Aroclor-1268-5	(5)	836674479	833098204	847063314	864357444	781067560	832452200	4
Decachlorobiphenyl		4074190590	4136234160	4241958160	4399826360	4004084600	4171258774	4
Tetrachloro-m-xylene		9139601460	8935957933	9311606620	9495189880	8951901000	9166851379	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM **Case No.:** P4495 **SAS No.:** P4495 **SDG NO.:** P4495

Instrument ID: ECD_O **Calibration Date(s):** 10/15/2024 10/16/2024
Calibration Times: 18:27 02:36

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO107184.D</u>		CF 750 = <u>PO107185.D</u>			
CF 500 = <u>PO107186.D</u>		CF 250 = <u>PO107187.D</u>		CF 050 = <u>PO107188.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	98519363	99934743	102532802	106601384	108740020	103265662 4
Aroclor-1016-2	(2)	140533795	142751421	144315942	146877552	130360060	140967754 5
Aroclor-1016-3	(3)	75819146	77111448	78799028	81783348	80215260	78745646 3
Aroclor-1016-4	(4)	61532960	63413097	65607722	69564368	70423600	66108349 6
Aroclor-1016-5	(5)	78878086	80970861	83522668	86940988	80356660	82133853 4
Aroclor-1260-1	(1)	150450531	152030172	155131756	161942364	156218200	155154605 3
Aroclor-1260-2	(2)	173752920	181526657	184104290	189736236	151457920	176115605 8
Aroclor-1260-3	(3)	168454604	170292340	172108108	175991544	151894120	167748143 6
Aroclor-1260-4	(4)	143637024	144895129	147115166	150514212	137242140	144680734 3
Aroclor-1260-5	(5)	340738400	339782780	336796716	340391952	289493480	329440666 7
Decachlorobiphenyl		2730622670	2745732653	2784792900	2852828640	2594005000	2741596373 3
Tetrachloro-m-xylene		3312014480	3338472613	3349065240	3230716080	2862214000	3218496483 6
Aroclor-1242-1	(1)	80390662	80447493	83937762	87266896	86674240	83743411 4
Aroclor-1242-2	(2)	114487494	111266404	117741916	120667112	114031640	115638913 3
Aroclor-1242-3	(3)	61817452	60461905	64396992	66865868	63624080	63433259 4
Aroclor-1242-4	(4)	60582529	60636219	64384594	68045260	66598220	64049364 5
Aroclor-1242-5	(5)	73763103	75675432	77215916	80736528	80226620	77523520 4
Decachlorobiphenyl		2686985320	2764418280	2764436780	2798541920	2665265400	2735929540 2
Tetrachloro-m-xylene		3315724680	3241948653	3341198040	3335574400	2920075400	3230904235 6
Aroclor-1248-1	(1)	60218805	63038185	64879120	66667236	62402400	63441149 4
Aroclor-1248-2	(2)	85173308	89108361	93061950	96541480	93540340	91485088 5
Aroclor-1248-3	(3)	89257070	93226611	97377878	100891904	96212540	95393201 5
Aroclor-1248-4	(4)	105934439	110533117	114760412	117772688	111691080	112138347 4
Aroclor-1248-5	(5)	102828085	105359652	109801380	114217964	119140640	110269544 6
Decachlorobiphenyl		2669365810	2682497173	2791873700	2836714480	2679938000	2732077833 3
Tetrachloro-m-xylene		3278417600	3389239440	3425836480	3362184040	2965722400	3284279992 6
Aroclor-1254-1	(1)	159855276	163854537	165548332	171717704	162964860	164788142 3
Aroclor-1254-2	(2)	138468840	142269335	144546922	151325972	148203140	144962842 3
Aroclor-1254-3	(3)	227061110	231385951	232213406	238575248	218867080	229620559 3
Aroclor-1254-4	(4)	128242906	130249965	131325350	135207044	122091880	129423429 4
Aroclor-1254-5	(5)	191904554	195153152	195378104	201188864	171636600	191052255 6
Decachlorobiphenyl		2726927150	2738190720	2773447120	2836815760	2598739800	2734824110 3
Tetrachloro-m-xylene		3347799560	3397335627	3354246660	3362236280	2930639600	3278451545 6
Aroclor-1268-1	(1)	419191400	409991492	413863584	410849472	368955300	404570250 5

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	389672996	380096764	383518208	378699872	334345600	373266688	6
Aroclor-1268-3	(3)	345284332	337828173	339535756	338755760	307004380	333681680	5
Aroclor-1268-4	(4)	129107993	127445005	127285362	129260796	113749480	125369727	5
Aroclor-1268-5	(5)	1049237500	1019311260	1017080130	996315476	856840280	987756929	8
Decachlorobiphenyl		4873889790	4739451973	4823511780	4890043960	4430545400	4751488581	4
Tetrachloro-m-xylene		3440286630	3329771373	3416104800	3331345720	2963806400	3296262985	6