

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

Continuing Calib Date: 10/28/2024 Initial Calibration Date(s): 10/28/2024 10/28/2024

Continuing Calib Time: 11:40 Initial Calibration Time(s): 09:34 10:46

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1221-1 (1)	4.57	4.57	4.47	4.67	0.00
Aroclor-1221-2 (2)	4.66	4.66	4.56	4.76	0.00
Aroclor-1221-3 (3)	4.73	4.73	4.63	4.83	0.00
Tetrachloro-m-xylene	4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl	10.06	10.06	9.96	10.16	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P4495 **SAS No.:** P4495 **SDG NO.:** P4495

Continuing Calib Date: 10/28/2024 **Initial Calibration Date(s):** 10/28/2024 10/28/2024

Continuing Calib Time: 11:40 **Initial Calibration Time(s):** 09:34 10:46

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1221-1 (1)	3.86	3.86	3.76	3.96	0.00
Aroclor-1221-2 (2)	3.94	3.94	3.84	4.04	0.00
Aroclor-1221-3 (3)	4.02	4.02	3.92	4.12	0.00
Tetrachloro-m-xylene	3.64	3.64	3.54	3.74	0.00
Decachlorobiphenyl	8.63	8.63	8.53	8.73	0.00