

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 10/25/2024

Lab Sample No.: AR1660CCC500 Data File : PO107370.D Time Analyzed: 09:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.523	5.423	5.623	527.200	500.000	5.4
Aroclor-1016-2	5.545	5.445	5.645	510.220	500.000	2.0
Aroclor-1016-3	5.608	5.507	5.707	494.160	500.000	-1.2
Aroclor-1016-4	5.705	5.604	5.804	514.300	500.000	2.9
Aroclor-1016-5	6.000	5.899	6.099	525.980	500.000	5.2
Aroclor-1260-1	7.126	7.026	7.226	489.010	500.000	-2.2
Aroclor-1260-2	7.383	7.282	7.482	517.380	500.000	3.5
Aroclor-1260-3	7.745	7.643	7.843	509.140	500.000	1.8
Aroclor-1260-4	7.969	7.868	8.068	518.330	500.000	3.7
Aroclor-1260-5	8.282	8.181	8.381	520.630	500.000	4.1
Decachlorobiphenyl	10.062	9.958	10.158	51.430	50.000	2.9
Tetrachloro-m-xylene	4.373	4.274	4.474	53.420	50.000	6.8

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 10/25/2024

Lab Sample No.: AR1660CCC500 Data File : PO107370.D Time Analyzed: 09:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.725	4.626	4.826	554.800	500.000	11.0
Aroclor-1016-2	4.744	4.645	4.845	567.110	500.000	13.4
Aroclor-1016-3	4.920	4.821	5.021	535.260	500.000	7.1
Aroclor-1016-4	4.961	4.862	5.062	509.110	500.000	1.8
Aroclor-1016-5	5.175	5.076	5.276	542.290	500.000	8.5
Aroclor-1260-1	6.205	6.107	6.307	541.100	500.000	8.2
Aroclor-1260-2	6.393	6.294	6.494	576.430	500.000	15.3
Aroclor-1260-3	6.546	6.448	6.648	552.590	500.000	10.5
Aroclor-1260-4	7.016	6.918	7.118	536.970	500.000	7.4
Aroclor-1260-5	7.257	7.159	7.359	571.380	500.000	14.3
Decachlorobiphenyl	8.636	8.539	8.739	52.070	50.000	4.1
Tetrachloro-m-xylene	3.644	3.545	3.745	55.350	50.000	10.7