

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 10/25/2024

Lab Sample No.: AR1660CCC500 Data File : PO107397.D Time Analyzed: 20:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.521	5.423	5.623	494.170	500.000	-1.2
Aroclor-1016-2	5.543	5.445	5.645	479.380	500.000	-4.1
Aroclor-1016-3	5.606	5.507	5.707	461.860	500.000	-7.6
Aroclor-1016-4	5.703	5.604	5.804	477.250	500.000	-4.6
Aroclor-1016-5	5.998	5.899	6.099	475.790	500.000	-4.8
Aroclor-1260-1	7.125	7.026	7.226	485.380	500.000	-2.9
Aroclor-1260-2	7.382	7.282	7.482	484.810	500.000	-3.0
Aroclor-1260-3	7.744	7.643	7.843	458.270	500.000	-8.3
Aroclor-1260-4	7.968	7.868	8.068	467.110	500.000	-6.6
Aroclor-1260-5	8.281	8.181	8.381	484.470	500.000	-3.1
Decachlorobiphenyl	10.061	9.958	10.158	50.240	50.000	0.5
Tetrachloro-m-xylene	4.372	4.274	4.474	52.530	50.000	5.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 10/25/2024

Lab Sample No.: AR1660CCC500 Data File : PO107397.D Time Analyzed: 20:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.725	4.626	4.826	537.990	500.000	7.6
Aroclor-1016-2	4.744	4.645	4.845	546.030	500.000	9.2
Aroclor-1016-3	4.920	4.821	5.021	516.060	500.000	3.2
Aroclor-1016-4	4.961	4.862	5.062	492.570	500.000	-1.5
Aroclor-1016-5	5.175	5.076	5.276	524.530	500.000	4.9
Aroclor-1260-1	6.206	6.107	6.307	496.750	500.000	-0.7
Aroclor-1260-2	6.393	6.294	6.494	527.570	500.000	5.5
Aroclor-1260-3	6.546	6.448	6.648	500.040	500.000	0.0
Aroclor-1260-4	7.017	6.918	7.118	488.580	500.000	-2.3
Aroclor-1260-5	7.258	7.159	7.359	504.820	500.000	1.0
Decachlorobiphenyl	8.637	8.539	8.739	47.900	50.000	-4.2
Tetrachloro-m-xylene	3.644	3.545	3.745	52.630	50.000	5.3