

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL05 Date Analyzed: 10/28/2024

Lab Sample No.: AR1221CCC500 Data File : PO107422.D Time Analyzed: 11:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1221-1	4.570	4.471	4.671	514.430	500.000	2.9
Aroclor-1221-2	4.655	4.556	4.756	510.810	500.000	2.2
Aroclor-1221-3	4.730	4.631	4.831	505.190	500.000	1.0
Decachlorobiphenyl	10.056	9.957	10.157	48.770	50.000	-2.5
Tetrachloro-m-xylene	4.371	4.271	4.471	51.410	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P4495 **SAS No.:** P4495 **SDG NO.:** P4495

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/28/2024 10/28/2024

Client Sample No.: CCAL05 **Date Analyzed:** 10/28/2024

Lab Sample No.: AR1221CCC500 **Data File :** PO107422.D **Time Analyzed:** 11:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1221-1	3.855	3.755	3.955	527.140	500.000	5.4
Aroclor-1221-2	3.940	3.840	4.040	513.260	500.000	2.7
Aroclor-1221-3	4.016	3.916	4.116	511.520	500.000	2.3
Decachlorobiphenyl	8.634	8.534	8.734	50.130	50.000	0.3
Tetrachloro-m-xylene	3.643	3.543	3.743	52.970	50.000	5.9