

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL07 Date Analyzed: 10/28/2024

Lab Sample No.: AR1221CCC500 Data File : PO107436.D Time Analyzed: 16:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1221-1	4.570	4.471	4.671	520.880	500.000	4.2
Aroclor-1221-2	4.655	4.556	4.756	510.950	500.000	2.2
Aroclor-1221-3	4.730	4.631	4.831	507.540	500.000	1.5
Decachlorobiphenyl	10.056	9.957	10.157	49.600	50.000	-0.8
Tetrachloro-m-xylene	4.371	4.271	4.471	52.200	50.000	4.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL07 Date Analyzed: 10/28/2024

Lab Sample No.: AR1221CCC500 Data File : PO107436.D Time Analyzed: 16:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1221-1	3.855	3.755	3.955	536.640	500.000	7.3
Aroclor-1221-2	3.941	3.840	4.040	516.940	500.000	3.4
Aroclor-1221-3	4.016	3.916	4.116	520.110	500.000	4.0
Decachlorobiphenyl	8.634	8.534	8.734	51.110	50.000	2.2
Tetrachloro-m-xylene	3.643	3.543	3.743	53.610	50.000	7.2