

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No. (PEM): PEM - PL092653.D Date Analyzed: 10/28/2024

Lab Sample No.(PEM): PEM Time Analyzed: 14:16

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.059	8.960	9.160	19.970	20.000	-0.2
Tetrachloro-m-xylene	3.546	3.500	3.600	19.290	20.000	-3.6
alpha-BHC	4.001	3.950	4.050	9.920	10.000	-0.8
beta-BHC	4.531	4.480	4.580	10.060	10.000	0.6
gamma-BHC (Lindane)	4.334	4.280	4.380	9.660	10.000	-3.4
Endrin	6.580	6.510	6.650	41.060	50.000	-17.9
4,4'-DDT	7.030	6.960	7.100	88.060	100.000	-11.9
Methoxychlor	7.505	7.430	7.580	204.090	250.000	-18.4

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No. (PEM): PEM - PL092653.D Date Analyzed: 10/28/2024

Lab Sample No.(PEM): PEM Time Analyzed: 14:16

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.918	7.820	8.020	19.080	20.000	-4.6
Tetrachloro-m-xylene	2.778	2.730	2.830	18.500	20.000	-7.5
alpha-BHC	3.281	3.230	3.330	8.630	10.000	-13.7
beta-BHC	3.911	3.860	3.960	9.760	10.000	-2.4
gamma-BHC (Lindane)	3.611	3.560	3.660	8.390	10.000	-16.1
Endrin	5.643	5.570	5.710	44.130	50.000	-11.7
4,4'-DDT	6.042	5.970	6.110	98.070	100.000	-1.9
Methoxychlor	6.616	6.550	6.690	225.800	250.000	-9.7