

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4495

Client: Chemtech Consulting Group

Analytical Method: 8081B Datafile : PL092641.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB164398BS	alpha-BHC	16.66	16.2	ug/kg	97				84	123	
	beta-BHC	16.66	16.7	ug/kg	100				82	123	
	delta-BHC	16.66	16.2	ug/kg	97				83	126	
	gamma-BHC (Lindane)	16.66	16.4	ug/kg	98				83	125	
	Heptachlor	16.66	16.7	ug/kg	100				83	122	
	Aldrin	16.66	16.3	ug/kg	98				82	124	
	Heptachlor epoxide	16.66	17.0	ug/kg	102				83	120	
	Endosulfan I	16.66	17.2	ug/kg	103				81	124	
	Dieldrin	16.66	16.4	ug/kg	98				85	121	
	4,4'-DDE	16.66	16.8	ug/kg	101				81	123	
	Endrin	16.66	16.1	ug/kg	97				76	130	
	Endosulfan II	16.66	16.3	ug/kg	98				80	125	
	4,4'-DDD	16.66	16.6	ug/kg	100				80	131	
	Endosulfan sulfate	16.66	16.3	ug/kg	98				81	122	
	4,4'-DDT	16.66	16.1	ug/kg	97				70	129	
	Methoxychlor	16.66	17.8	ug/kg	107				60	119	
	Endrin ketone	16.66	17.2	ug/kg	103				77	132	
	Endrin aldehyde	16.66	16.6	ug/kg	100				79	124	
	alpha-Chlordane	16.66	16.7	ug/kg	100				84	120	
	gamma-Chlordane	16.66	16.4	ug/kg	98				83	122	