

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

Continuing Calib Date: 10/31/2024 Initial Calibration Date(s): 10/28/2024 10/28/2024

Continuing Calib Time: 13:03 Initial Calibration Time(s): 16:56 17:50

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Toxaphene-1	(1)	6.24	6.24	6.14	6.34	0.00
Toxaphene-2	(2)	6.45	6.44	6.34	6.54	-0.01
Toxaphene-3	(3)	7.06	7.06	6.96	7.16	0.00
Toxaphene-4	(4)	7.16	7.15	7.05	7.25	-0.01
Toxaphene-5	(5)	7.94	7.93	7.83	8.03	-0.01
Decachlorobiphenyl		9.06	9.05	8.95	9.15	-0.01
Tetrachloro-m-xylene		3.55	3.54	3.44	3.64	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

Continuing Calib Date: 10/31/2024 Initial Calibration Date(s): 10/28/2024 10/28/2024

Continuing Calib Time: 13:03 Initial Calibration Time(s): 16:56 17:50

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Toxaphene-1 (1)	5.01	5.01	4.91	5.11	0.00
Toxaphene-2 (2)	5.33	5.33	5.23	5.43	0.00
Toxaphene-3 (3)	6.61	6.61	6.51	6.71	0.00
Toxaphene-4 (4)	6.74	6.73	6.63	6.83	-0.01
Toxaphene-5 (5)	7.05	7.05	6.95	7.15	0.00
Decachlorobiphenyl	7.92	7.92	7.82	8.02	0.00
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00