

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 10/31/2024

Lab Sample No.: PTOXCCC500 Data File : PL092756.D Time Analyzed: 13:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.060	8.954	9.154	47.840	50.000	-4.3
Tetrachloro-m-xylene	3.545	3.440	3.640	49.410	50.000	-1.2
Toxaphene-1	6.241	6.137	6.337	522.800	500.000	4.6
Toxaphene-2	6.446	6.341	6.541	582.480	500.000	16.5
Toxaphene-3	7.064	6.958	7.158	472.340	500.000	-5.5
Toxaphene-4	7.155	7.049	7.249	470.160	500.000	-6.0
Toxaphene-5	7.940	7.834	8.034	463.440	500.000	-7.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 10/31/2024

Lab Sample No.: PTOXCCC500 Data File : PL092756.D Time Analyzed: 13:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.918	7.816	8.016	48.610	50.000	-2.8
Tetrachloro-m-xylene	2.779	2.678	2.878	50.310	50.000	0.6
Toxaphene-1	5.008	4.907	5.107	598.250	500.000	19.7
Toxaphene-2	5.333	5.232	5.432	511.660	500.000	2.3
Toxaphene-3	6.607	6.505	6.705	467.970	500.000	-6.4
Toxaphene-4	6.735	6.633	6.833	502.400	500.000	0.5
Toxaphene-5	7.048	6.945	7.145	474.260	500.000	-5.1