

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 10/31/2024

Lab Sample No.: PTOXCCC500 Data File : PL092770.D Time Analyzed: 18:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.059	8.954	9.154	49.030	50.000	-1.9
Tetrachloro-m-xylene	3.542	3.440	3.640	49.910	50.000	-0.2
Toxaphene-1	6.239	6.137	6.337	545.710	500.000	9.1
Toxaphene-2	6.444	6.341	6.541	562.140	500.000	12.4
Toxaphene-3	7.062	6.958	7.158	475.650	500.000	-4.9
Toxaphene-4	7.152	7.049	7.249	481.670	500.000	-3.7
Toxaphene-5	7.937	7.834	8.034	468.130	500.000	-6.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 10/31/2024

Lab Sample No.: PTOXCCC500 Data File : PL092770.D Time Analyzed: 18:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.918	7.816	8.016	49.790	50.000	-0.4
Tetrachloro-m-xylene	2.778	2.678	2.878	50.570	50.000	1.1
Toxaphene-1	5.007	4.907	5.107	590.520	500.000	18.1
Toxaphene-2	5.333	5.232	5.432	519.450	500.000	3.9
Toxaphene-3	6.606	6.505	6.705	470.010	500.000	-6.0
Toxaphene-4	6.734	6.633	6.833	497.860	500.000	-0.4
Toxaphene-5	7.047	6.945	7.145	478.080	500.000	-4.4