

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111324\
 Data File : BF140341.D
 Acq On : 13 Nov 2024 13:19
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF111324

Quant Time: Nov 13 14:41:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 14:40:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.522	0.517	1.0	102	0.00
3	Pyridine	1.283	1.305	-1.7	101	0.00
4	n-Nitrosodimethylamine	0.649	0.647	0.3	100	0.00
5 S	2-Fluorophenol	1.171	1.140	2.6	101	0.00
6	Aniline	1.381	1.433	-3.8	101	0.00
7 S	Phenol-d6	1.587	1.572	0.9	100	0.00
8	2-Chlorophenol	1.258	1.245	1.0	100	0.00
9	Benzaldehyde	0.951	0.888	6.6	102	0.00
10 C	Phenol	1.674	1.685	-0.7	100	0.00
11	bis(2-Chloroethyl)ether	1.268	1.248	1.6	101	0.00
12	1,3-Dichlorobenzene	1.388	1.359	2.1	101	0.00
13 C	1,4-Dichlorobenzene	1.408	1.370	2.7	100	0.00
14	1,2-Dichlorobenzene	1.307	1.273	2.6	100	0.00
15	Benzyl Alcohol	1.143	1.189	-4.0	100	0.00
16	2,2'-oxybis(1-Chloropropane	1.752	1.765	-0.7	99	0.00
17	2-Methylphenol	1.035	1.041	-0.6	100	0.00
18	Hexachloroethane	0.520	0.509	2.1	101	0.00
19 P	n-Nitroso-di-n-propylamine	0.959	0.960	-0.1	100	0.00
20	3+4-Methylphenols	1.294	1.328	-2.6	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
22	Acetophenone	0.465	0.451	3.0	100	0.00
23 S	Nitrobenzene-d5	0.384	0.377	1.8	100	0.00
24	Nitrobenzene	0.400	0.393	1.8	100	0.00
25	Isophorone	0.680	0.673	1.0	100	0.00
26 C	2-Nitrophenol	0.177	0.181	-2.3	101	0.00
27	2,4-Dimethylphenol	0.227	0.230	-1.3	103	0.00
28	bis(2-Chloroethoxy)methane	0.417	0.409	1.9	100	0.00
29 C	2,4-Dichlorophenol	0.281	0.277	1.4	101	0.00
30	1,2,4-Trichlorobenzene	0.312	0.303	2.9	100	0.00
31	Naphthalene	1.015	0.986	2.9	100	0.00
32	Benzoic acid	0.200	0.213	-6.5	100	0.00
33	4-Chloroaniline	0.353	0.344	2.5	101	0.00
34 C	Hexachlorobutadiene	0.199	0.193	3.0	100	0.00
35	Caprolactam	0.093	0.094	-1.1	102	0.00
36 C	4-Chloro-3-methylphenol	0.311	0.314	-1.0	101	0.00
37	2-Methylnaphthalene	0.651	0.635	2.5	100	0.00
38	1-Methylnaphthalene	0.637	0.620	2.7	99	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.553	0.534	3.4	100	0.00
41 P	Hexachlorocyclopentadiene	0.142	0.136	4.2	97	0.00
42 S	2,4,6-Tribromophenol	0.199	0.195	2.0	101	0.00
43 C	2,4,6-Trichlorophenol	0.363	0.361	0.6	100	0.00
44	2,4,5-Trichlorophenol	0.397	0.382	3.8	98	0.00
45 S	2-Fluorobiphenyl	1.244	1.177	5.4	100	0.00
46	1,1'-Biphenyl	1.455	1.421	2.3	101	0.00
47	2-Chloronaphthalene	1.108	1.068	3.6	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111324\
 Data File : BF140341.D
 Acq On : 13 Nov 2024 13:19
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF111324

Quant Time: Nov 13 14:41:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 14:40:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.368	0.366	0.5	100	0.00
49	Acenaphthylene	1.677	1.633	2.6	99	0.00
50	Dimethylphthalate	1.304	1.269	2.7	100	0.00
51	2,6-Dinitrotoluene	0.297	0.297	0.0	99	0.00
52 C	Acenaphthene	1.133	1.108	2.2	101	0.00
53	3-Nitroaniline	0.312	0.311	0.3	100	0.00
54 P	2,4-Dinitrophenol	0.153	0.154	-0.7	104	0.00
55	Dibenzofuran	1.603	1.575	1.7	101	0.00
56 P	4-Nitrophenol	0.228	0.238	-4.4	100	0.00
57	2,4-Dinitrotoluene	0.391	0.394	-0.8	101	0.00
58	Fluorene	1.267	1.218	3.9	101	0.00
59	2,3,4,6-Tetrachlorophenol	0.313	0.310	1.0	97	0.00
60	Diethylphthalate	1.333	1.298	2.6	100	0.00
61	4-Chlorophenyl-phenylether	0.625	0.601	3.8	99	0.00
62	4-Nitroaniline	0.319	0.316	0.9	98	0.00
63	Azobenzene	1.317	1.306	0.8	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.108	0.111	-2.8	102	0.00
66 c	n-Nitrosodiphenylamine	0.578	0.562	2.8	101	0.00
67	4-Bromophenyl-phenylether	0.200	0.195	2.5	101	0.00
68	Hexachlorobenzene	0.225	0.220	2.2	101	0.00
69	Atrazine	0.175	0.143	18.3	99	0.00
70 C	Pentachlorophenol	0.121	0.126	-4.1	100	0.00
71	Phenanthrene	0.940	0.893	5.0	99	0.00
72	Anthracene	0.921	0.883	4.1	100	0.00
73	Carbazole	0.909	0.879	3.3	101	0.00
74	Di-n-butylphthalate	1.091	1.075	1.5	100	0.00
75 C	Fluoranthene	1.054	1.007	4.5	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	95	0.01
77	Benzydine	0.768	0.746	2.9	108	0.00
78	Pyrene	1.711	1.739	-1.6	98	0.00
79 S	Terphenyl-d14	1.152	1.164	-1.0	98	0.00
80	Butylbenzylphthalate	0.657	0.699	-6.4	96	0.00
81	Benzo(a)anthracene	1.314	1.292	1.7	97	0.01
82	3,3'-Dichlorobenzidine	0.418	0.416	0.5	98	0.01
83	Chrysene	1.199	1.186	1.1	96	0.00
84	Bis(2-ethylhexyl)phthalate	0.877	0.926	-5.6	95	0.01
85 c	Di-n-octyl phthalate	1.235	1.249	-1.1	92	0.02
86 I	Perylene-d12	1.000	1.000	0.0	99	0.02
87	Indeno(1,2,3-cd)pyrene	1.235	1.278	-3.5	101	0.02
88	Benzo(b)fluoranthene	1.308	1.215	7.1	95	0.02
89	Benzo(k)fluoranthene	1.069	1.085	-1.5	101	0.02
90 C	Benzo(a)pyrene	1.016	0.999	1.7	98	0.02
91	Dibenzo(a,h)anthracene	1.021	1.069	-4.7	102	0.02
92	Benzo(g,h,i)perylene	1.044	1.061	-1.6	99	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111324\
Data File : BF140341.D
Acq On : 13 Nov 2024 13:19
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF111324

Quant Time: Nov 13 14:41:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 14:40:06 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0