

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111324\
 Data File : BF140341.D
 Acq On : 13 Nov 2024 13:19
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF111324

Quant Time: Nov 13 14:41:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 14:40:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	39.597	1.0	102	0.00
3	Pyridine	40.000	40.655	-1.6	101	0.00
4	n-Nitrosodimethylamine	40.000	39.908	0.2	100	0.00
5 S	2-Fluorophenol	80.000	77.878	2.7	101	0.00
6	Aniline	40.000	41.531	-3.8	101	0.00
7 S	Phenol-d6	80.000	79.248	0.9	100	0.00
8	2-Chlorophenol	40.000	39.588	1.0	100	0.00
9	Benzaldehyde	40.000	37.337	6.7	102	0.00
10 C	Phenol	40.000	40.259	-0.6	100	0.00
11	bis(2-Chloroethyl)ether	40.000	39.367	1.6	101	0.00
12	1,3-Dichlorobenzene	40.000	39.148	2.1	101	0.00
13 C	1,4-Dichlorobenzene	40.000	38.930	2.7	100	0.00
14	1,2-Dichlorobenzene	40.000	38.953	2.6	100	0.00
15	Benzyl Alcohol	40.000	41.596	-4.0	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.304	-0.8	99	0.00
17	2-Methylphenol	40.000	40.228	-0.6	100	0.00
18	Hexachloroethane	40.000	39.149	2.1	101	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.079	-0.2	100	0.00
20	3+4-Methylphenols	40.000	41.028	-2.6	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	38.782	3.0	100	0.00
23 S	Nitrobenzene-d5	80.000	78.511	1.9	100	0.00
24	Nitrobenzene	40.000	39.284	1.8	100	0.00
25	Isophorone	40.000	39.558	1.1	100	0.00
26 C	2-Nitrophenol	40.000	40.850	-2.1	101	0.00
27	2,4-Dimethylphenol	40.000	40.536	-1.3	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.258	1.9	100	0.00
29 C	2,4-Dichlorophenol	40.000	39.545	1.1	101	0.00
30	1,2,4-Trichlorobenzene	40.000	38.843	2.9	100	0.00
31	Naphthalene	40.000	38.889	2.8	100	0.00
32	Benzoic acid	40.000	42.635	-6.6	100	0.00
33	4-Chloroaniline	40.000	38.996	2.5	101	0.00
34 C	Hexachlorobutadiene	40.000	38.804	3.0	100	0.00
35	Caprolactam	40.000	40.345	-0.9	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.356	-0.9	101	0.00
37	2-Methylnaphthalene	40.000	39.072	2.3	100	0.00
38	1-Methylnaphthalene	40.000	38.933	2.7	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.657	3.4	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.267	4.3	97	0.00
42 S	2,4,6-Tribromophenol	80.000	78.617	1.7	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.769	0.6	100	0.00
44	2,4,5-Trichlorophenol	40.000	38.542	3.6	98	0.00
45 S	2-Fluorobiphenyl	80.000	75.707	5.4	100	0.00
46	1,1'-Biphenyl	40.000	39.068	2.3	101	0.00
47	2-Chloronaphthalene	40.000	38.543	3.6	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111324\
 Data File : BF140341.D
 Acq On : 13 Nov 2024 13:19
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF111324

Quant Time: Nov 13 14:41:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 14:40:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.821	0.4	100	0.00
49	Acenaphthylene	40.000	38.962	2.6	99	0.00
50	Dimethylphthalate	40.000	38.933	2.7	100	0.00
51	2,6-Dinitrotoluene	40.000	39.952	0.1	99	0.00
52 C	Acenaphthene	40.000	39.143	2.1	101	0.00
53	3-Nitroaniline	40.000	39.809	0.5	100	0.00
54 P	2,4-Dinitrophenol	40.000	40.153	-0.4	104	0.00
55	Dibenzofuran	40.000	39.280	1.8	101	0.00
56 P	4-Nitrophenol	40.000	41.782	-4.5	100	0.00
57	2,4-Dinitrotoluene	40.000	40.258	-0.6	101	0.00
58	Fluorene	40.000	38.457	3.9	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.601	1.0	97	0.00
60	Diethylphthalate	40.000	38.945	2.6	100	0.00
61	4-Chlorophenyl-phenylether	40.000	38.465	3.8	99	0.00
62	4-Nitroaniline	40.000	39.561	1.1	98	0.00
63	Azobenzene	40.000	39.649	0.9	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.144	-2.9	102	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.926	2.7	101	0.00
67	4-Bromophenyl-phenylether	40.000	39.059	2.4	101	0.00
68	Hexachlorobenzene	40.000	39.043	2.4	101	0.00
69	Atrazine	40.000	32.640	18.4	99	0.00
70 C	Pentachlorophenol	40.000	41.693	-4.2	100	0.00
71	Phenanthrene	40.000	38.024	4.9	99	0.00
72	Anthracene	40.000	38.362	4.1	100	0.00
73	Carbazole	40.000	38.676	3.3	101	0.00
74	Di-n-butylphthalate	40.000	39.391	1.5	100	0.00
75 C	Fluoranthene	40.000	38.196	4.5	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	95	0.01
77	Benzydine	40.000	38.866	2.8	108	0.00
78	Pyrene	40.000	40.667	-1.7	98	0.00
79 S	Terphenyl-d14	80.000	80.797	-1.0	98	0.00
80	Butylbenzylphthalate	40.000	42.560	-6.4	96	0.00
81	Benzo(a)anthracene	40.000	39.327	1.7	97	0.01
82	3,3'-Dichlorobenzidine	40.000	39.793	0.5	98	0.01
83	Chrysene	40.000	39.569	1.1	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.213	-5.5	95	0.01
85 c	Di-n-octyl phthalate	40.000	40.448	-1.1	92	0.02
86 I	Perylene-d12	20.000	20.000	0.0	99	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	41.371	-3.4	101	0.02
88	Benzo(b)fluoranthene	40.000	37.153	7.1	95	0.02
89	Benzo(k)fluoranthene	40.000	40.625	-1.6	101	0.02
90 C	Benzo(a)pyrene	40.000	39.337	1.7	98	0.02
91	Dibenzo(a,h)anthracene	40.000	41.871	-4.7	102	0.02
92	Benzo(g,h,i)perylene	40.000	40.635	-1.6	99	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111324\
Data File : BF140341.D
Acq On : 13 Nov 2024 13:19
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF111324

Quant Time: Nov 13 14:41:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 14:40:06 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0