

Surrogate Summary

SW-846

SDG No.: P4495

Client: Chemtech Consulting Group

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4495-11	PT-BNA-SOIL	2-Fluorophenol	150	88.0	59		18	112
		Phenol-d6	150	91.3	61		15	107
		Nitrobenzene-d5	100	64.2	64		18	107
		2-Fluorobiphenyl	100	65.3	65		20	109
		2,4,6-Tribromophenol	150	99.2	66		10	116
		Terphenyl-d14	100	83.0	83		10	105
P4495-11DL	PT-BNA-SOILDL	2-Fluorophenol	150	82.4	55		18	112
		Phenol-d6	150	84.0	56		15	107
		Nitrobenzene-d5	100	57.4	57		18	107
		2-Fluorobiphenyl	100	64.8	65		20	109
		2,4,6-Tribromophenol	150	91.7	61		10	116
		Terphenyl-d14	100	78.4	78		10	105
P4547-05MS	BP-F-20MS	2-Fluorophenol	150	90.8	61		18	112
		Phenol-d6	150	91.2	61		15	107
		Nitrobenzene-d5	100	66.4	66		18	107
		2-Fluorobiphenyl	100	61.5	61		20	109
		2,4,6-Tribromophenol	150	106	71		10	116
		Terphenyl-d14	100	68.7	69		10	105
P4547-05MSD	BP-F-20MSD	2-Fluorophenol	150	90.8	61		18	112
		Phenol-d6	150	91.9	61		15	107
		Nitrobenzene-d5	100	66.5	66		18	107
		2-Fluorobiphenyl	100	61.6	62		20	109
		2,4,6-Tribromophenol	150	109	73		10	116
		Terphenyl-d14	100	70.4	70		10	105
PB164401BL	PB164401BL	2-Fluorophenol	150	129	86		18	112
		Phenol-d6	150	126	84		15	107
		Nitrobenzene-d5	100	87.1	87		18	107
		2-Fluorobiphenyl	100	85.2	85		20	109
		2,4,6-Tribromophenol	150	121	80		10	116
		Terphenyl-d14	100	82.5	82		10	105
PB164401BS	PB164401BS	2-Fluorophenol	150	115	77		18	112
		Phenol-d6	150	112	75		15	107
		Nitrobenzene-d5	100	85.7	86		18	107
		2-Fluorobiphenyl	100	83.9	84		20	109
		2,4,6-Tribromophenol	150	135	90		10	116
		Terphenyl-d14	100	87.9	88		10	105