

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111424\
 Data File : BN035082.D
 Acq On : 14 Nov 2024 10:12
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 14 11:23:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 17:18:14 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	91	0.00
2	1,4-Dioxane	0.400	0.390	2.5	96	0.00
3	n-Nitrosodimethylamine	0.400	0.354	11.5	84	0.00
4 S	2-Fluorophenol	0.400	0.361	9.8	88	0.00
5 S	Phenol-d6	0.400	0.372	7.0	93	0.00
6	bis(2-Chloroethyl)ether	0.400	0.370	7.5	90	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	93	0.00
8 S	Nitrobenzene-d5	0.400	0.354	11.5	89	-0.01
9	Naphthalene	0.400	0.369	7.8	92	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.368	8.0	92	0.00
12	2-Methylnaphthalene	0.400	0.368	8.0	92	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	93	0.00
14 S	2,4,6-Tribromophenol	0.400	0.318	20.5	89	0.00
15 S	2-Fluorobiphenyl	0.400	0.358	10.5	93	0.00
16	Acenaphthylene	0.400	0.346	13.5	93	0.00
17	Acenaphthene	0.400	0.355	11.3	93	0.00
18	Fluorene	0.400	0.354	11.5	93	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	95	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.338	15.5	92	0.00
21	4-Bromophenyl-phenylether	0.400	0.376	6.0	92	0.00
22	Hexachlorobenzene	0.400	0.378	5.5	94	0.00
23	Atrazine	0.400	0.341	14.8	85	0.00
24	Pentachlorophenol	0.400	0.337	15.8	95	0.00
25	Phenanthrene	0.400	0.380	5.0	94	0.00
26	Anthracene	0.400	0.370	7.5	92	0.00
27 SURR	Fluoranthene-d10	0.400	0.369	7.8	92	0.00
28	Fluoranthene	0.400	0.372	7.0	92	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	90	0.00
30	Pyrene	0.400	0.394	1.5	92	0.00
31 S	Terphenyl-d14	0.400	0.387	3.3	90	0.00
32	Benzo(a)anthracene	0.400	0.378	5.5	89	0.00
33	Chrysene	0.400	0.389	2.8	91	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.318	20.5	81	0.00
35 I	Perylene-d12	0.400	0.400	0.0	86	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.372	7.0	84	0.00
37	Benzo(b)fluoranthene	0.400	0.385	3.8	87	0.00
38	Benzo(k)fluoranthene	0.400	0.402	-0.5	91	0.00
39 C	Benzo(a)pyrene	0.400	0.379	5.3	86	0.00
40	Dibenzo(a,h)anthracene	0.400	0.372	7.0	84	0.00
41	Benzo(g,h,i)perylene	0.400	0.359	10.3	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4