

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN111324

Review By	yogesh	Review On	11/14/2024 3:07:49 AM
Supervise By	mohammad	Supervise On	11/14/2024 3:22:04 AM
SubDirectory	BN111324	HP Acquire Method	BNA_N, 8270_HP Processing Method BN111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597
CCC	SP6601
Internal Standard/PEM	SP6527
ICV/I.BLK	SP6548
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN035061.D	13 Nov 2024 12:01		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN035062.D	13 Nov 2024 12:40		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN035063.D	13 Nov 2024 13:16		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN035064.D	13 Nov 2024 13:52		RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN035065.D	13 Nov 2024 14:28		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN035066.D	13 Nov 2024 15:04		RC/JU	Ok,M
7	SSTDICC3.2	SSTDICC3.2	BN035067.D	13 Nov 2024 15:39		RC/JU	Ok,M
8	SSTDICC5.0	SSTDICC5.0	BN035068.D	13 Nov 2024 16:15		RC/JU	Ok
9	SSTDICV0.4	ICVBN111324	BN035069.D	13 Nov 2024 16:51		RC/JU	Ok,M
10	PB164785BL	PB164785BL	BN035070.D	13 Nov 2024 17:27		RC/JU	Ok
11	PB164785BS	PB164785BS	BN035071.D	13 Nov 2024 18:03		RC/JU	Ok
12	PB164785BSD	PB164785BSD	BN035072.D	13 Nov 2024 18:38		RC/JU	Ok
13	P4765-02	DSN002	BN035073.D	13 Nov 2024 19:14		RC/JU	Ok
14	P4765-06	DSN003	BN035074.D	13 Nov 2024 19:50		RC/JU	Ok
15	P4767-02	TOWER-2	BN035075.D	13 Nov 2024 20:26		RC/JU	Ok,M
16	P4773-01	RW7-SP100-20241107	BN035076.D	13 Nov 2024 21:01	Need 2X Dilution	RC/JU	Dilution
17	P4773-02	RW7-SP201-20241107	BN035077.D	13 Nov 2024 21:37		RC/JU	Ok
18	P4774-06	BP-VPB-190-GW-638-6	BN035078.D	13 Nov 2024 22:13		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN111324

Review By	yogesh	Review On	11/14/2024 3:07:49 AM
Supervise By	mohammad	Supervise On	11/14/2024 3:22:04 AM
SubDirectory	BN111324	HP Acquire Method	BNA_N, 8270_HP Processing Method BN111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597
CCC	SP6601
Internal Standard/PEM	SP6527
ICV/I.BLK	SP6548
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P4774-02	VPB190-HYD-2024110	BN035079.D	13 Nov 2024 22:49	Surrogate Fail	RC/JU	ReRun
20	SSTDCCC0.4	SSTDCCC0.4EC	BN035080.D	13 Nov 2024 23:24		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN111424

Review By	yogesh	Review On	11/15/2024 5:16:11 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:46 AM
SubDirectory	BN111424	HP Acquire Method	BNA_N, 8270_HP Processing Method BN111324
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597		
CCC	SP6601		
Internal Standard/PEM	SP6527		
ICV/I.BLK	SP6548		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN035081.D	14 Nov 2024 08:53		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN035082.D	14 Nov 2024 10:12		RC/JU	Ok
3	PB164402BL	PB164402BL	BN035083.D	14 Nov 2024 10:48		RC/JU	Ok
4	P4495-11	PT-BNA-SOIL	BN035084.D	14 Nov 2024 11:24	PT Sample, Need 50X Dilution	RC/JU	Dilution
5	P4495-11DL	PT-BNA-SOILDL	BN035085.D	14 Nov 2024 12:31		RC/JU	Ok
6	P4495-13	PT-PAH-SOIL	BN035086.D	14 Nov 2024 13:22	PT Sample, Need 5X Dilution	RC/JU	Dilution
7	P4495-13DL	PT-PAH-SOILDL	BN035087.D	14 Nov 2024 14:10		RC/JU	Ok
8	P3845-21	RR-PAH-WP	BN035088.D	14 Nov 2024 14:59	PT Sample Analyzed for confirmation, Need 10X Dilution	RC/JU	Dilution
9	P3845-21DL	RR-PAH-WPDL	BN035089.D	14 Nov 2024 15:58	PT Sample Analyzed for confirmation	RC/JU	Ok,M
10	P4774-02RE	VPB190-HYD-20241107	BN035090.D	14 Nov 2024 16:35	Surrogate Fail	RC/JU	Confirms
11	P4773-04	RW7-SP303-20241107	BN035091.D	14 Nov 2024 17:11		RC/JU	Ok
12	P4773-03	RW7-SP302-20241107	BN035092.D	14 Nov 2024 17:47		RC/JU	Ok
13	P4773-01DL	RW7-SP100-20241107	BN035093.D	14 Nov 2024 18:23		RC/JU	Ok
14	PB164402BS	PB164402BS	BN035094.D	14 Nov 2024 18:59		RC/JU	Ok
15	SSTDCCC0.4	SSTDCCC0.4EC	BN035095.D	14 Nov 2024 19:35		RC/JU	Ok

M : Manual Integration