

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4495

Client: Chemtech Consulting Group

Analytical Method: 8270-Modified DataFile: BN035094.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB164402BS	Naphthalene	13.3	12.0	ug/Kg	90				63	120	
	2-Methylnaphthalene	13.3	12.0	ug/Kg	90				54	111	
	Acenaphthylene	13.3	12.1	ug/Kg	91				61	110	
	Acenaphthene	13.3	11.6	ug/Kg	87				65	112	
	Fluorene	13.3	11.2	ug/Kg	84				61	115	
	Phenanthrene	13.3	12.3	ug/Kg	92				61	120	
	Anthracene	13.3	12.5	ug/Kg	94				60	115	
	Fluoranthene	13.3	10.7	ug/Kg	80				56	124	
	Pyrene	13.3	13.6	ug/Kg	102				58	125	
	Benzo(a)anthracene	13.3	12.0	ug/Kg	90				45	129	
	Chrysene	13.3	12.6	ug/Kg	95				62	123	
	Benzo(b)fluoranthene	13.3	13.0	ug/Kg	98				47	123	
	Benzo(k)fluoranthene	13.3	12.8	ug/Kg	96				69	123	
	Benzo(a)pyrene	13.3	13.0	ug/Kg	98				53	123	
	Indeno(1,2,3-cd)pyrene	13.3	11.2	ug/Kg	84				62	131	
	Dibenz(a,h)anthracene	13.3	11.2	ug/Kg	84				57	120	
	Benzo(g,h,i)perylene	13.3	10.0	ug/Kg	75				66	121	