

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 11/14/2024

Lab File ID: BN035082.D Time Analyzed: 10:12

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2551	7.553	6362	10.32	5037	14.19
UPPER LIMIT	5102	8.053	12724	10.819	10074	14.69
LOWER LIMIT	1275.5	7.053	3181	9.819	2518.5	13.69
EPA SAMPLE NO.						
01 PB164402BL	2707	7.55	6301	10.32	4578	14.19
02 PT-PAH-SOIL	3754	7.55	10271	10.31	7900	14.19
03 PT-PAH-SOILDL	3588	7.55	9962	10.32	7625	14.19
04 PB164402BS	2568	7.55	6459	10.32	4848	14.19

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.