

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG No.: P4495
Instrument ID: BNA_F Calibration Date(s): 10/18/2024 10/18/2024
Calibration Time(s): 10:27 13:46

LAB FILE ID:		RRF2.5 = BF139844.D		RRF005 = BF139845.D		RRF010 = BF139846.D		
		RRF020 = BF139847.D		RRF040 = BF139848.D		RRF050 = BF139849.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.465	1.398	1.331	1.278	1.179	1.278	10.1
Phenol-d6		1.900	1.818	1.709	1.647	1.538	1.655	10.1
Nitrobenzene-d5		0.365	0.365	0.372	0.371	0.354	0.361	2.9
2-Fluorobiphenyl		1.512	1.396	1.280	1.162	1.088	1.210	16.0
2,4,6-Tribromophenol		0.201	0.196	0.193	0.188	0.179	0.187	5.5
Atrazine		0.189	0.175	0.156	0.175	0.135	0.162	11.4
Terphenyl-d14		1.381	1.335	1.340	1.244	1.154	1.227	11.0