

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF101824

Review By	yogesh	Review On	10/21/2024 6:34:01 AM
Supervise By	mohammad	Supervise On	10/21/2024 6:38:35 AM
SubDirectory	BF101824	HP Acquire Method	BNA_F
		HP Processing Method	bf101824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12322,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139843.D	18 Oct 2024 09:22		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF139844.D	18 Oct 2024 10:27		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF139845.D	18 Oct 2024 10:55	Compound #9,32,54,65,85 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF139846.D	18 Oct 2024 11:23		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF139847.D	18 Oct 2024 11:52	Compound #54 Kept on LR	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF139848.D	18 Oct 2024 12:20	The Calibration is Good For 8270 DOD Except com#77 and good for 625.1 Method	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF139849.D	18 Oct 2024 12:49	Com#77(Benzidine) Failed in the calibration for both DOD and NON-DOD	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF139850.D	18 Oct 2024 13:17		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF139851.D	18 Oct 2024 13:46		RC/JU	Ok,M
10	SSTDICV040	SSTDICV040	BF139852.D	18 Oct 2024 14:19		RC/JU	Ok,M
11	PB164211BL	PB164211BL	BF139853.D	18 Oct 2024 14:48		RC/JU	Ok
12	P4405-01	MH-121	BF139854.D	18 Oct 2024 15:21		RC/JU	Ok
13	P4431-01	72-11934	BF139855.D	18 Oct 2024 15:50		RC/JU	Ok,M
14	P4421-01	EO-02-101624	BF139856.D	18 Oct 2024 16:18		RC/JU	Ok,M
15	P4422-01	EO-01-101624	BF139857.D	18 Oct 2024 16:46		RC/JU	Ok
16	P4425-01	TP-1	BF139858.D	18 Oct 2024 17:15		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF101824

Review By	yogesh	Review On	10/21/2024 6:34:01 AM		
Supervise By	mohammad	Supervise On	10/21/2024 6:38:35 AM		
SubDirectory	BF101824	HP Acquire Method	BNA_F	HP Processing Method	bf101824
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12322,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

17	P4425-03	TP-2	BF139859.D	18 Oct 2024 17:44		RC/JU	Ok
18	P4425-05	TP-3	BF139860.D	18 Oct 2024 18:12		RC/JU	Ok
19	P4425-07	TP-4	BF139861.D	18 Oct 2024 18:41	Internal Standrad Fail	RC/JU	ReRun
20	P4425-09	TP-5	BF139862.D	18 Oct 2024 19:09	Internal Standrad Fail	RC/JU	ReRun
21	P4426-03	PAD-2	BF139863.D	18 Oct 2024 19:37	Internal Standrad Fail	RC/JU	ReRun
22	P4426-07	PAD-4	BF139864.D	18 Oct 2024 20:05	Internal Standrad Fail	RC/JU	ReRun
23	P4426-17	PAD-9	BF139865.D	18 Oct 2024 20:34	Internal Standard Fail	RC/JU	ReRun
24	P4426-11	PAD-6	BF139866.D	18 Oct 2024 21:02	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF102624

Review By	yogesh	Review On	10/29/2024 1:34:36 AM
Supervise By	mohammad	Supervise On	10/29/2024 1:38:32 AM
SubDirectory	BF102624	HP Acquire Method	BNA_F
		HP Processing Method	bf101824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12323,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140049.D	26 Oct 2024 10:10		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140050.D	26 Oct 2024 10:38		RC/JU	Ok,M
3	PB164401BL	PB164401BL	BF140051.D	26 Oct 2024 11:06		RC/JU	Ok
4	PB164401BS	PB164401BS	BF140052.D	26 Oct 2024 11:34		RC/JU	Ok,M
5	P4508-09	BP-F22	BF140053.D	26 Oct 2024 12:06		RC/JU	Ok
6	P4508-09MS	BP-F22MS	BF140054.D	26 Oct 2024 12:34		RC/JU	Ok,M
7	P4508-09MSD	BP-F22MSD	BF140055.D	26 Oct 2024 13:02		RC/JU	Ok,M
8	P4547-05	BP-F-20	BF140056.D	26 Oct 2024 13:30		RC/JU	Ok
9	P4547-05MS	BP-F-20MS	BF140057.D	26 Oct 2024 13:58		RC/JU	Ok,M
10	P4547-05MSD	BP-F-20MSD	BF140058.D	26 Oct 2024 14:27		RC/JU	Ok,M
11	P4508-05	BP-F23	BF140059.D	26 Oct 2024 14:55		RC/JU	Ok
12	P4517-07	FOREST-ST-CO	BF140060.D	26 Oct 2024 15:23		RC/JU	Ok
13	P4487-08	BP-B27	BF140061.D	26 Oct 2024 15:51		RC/JU	Ok
14	P4460-04	WB-303-BOT	BF140062.D	26 Oct 2024 16:20		RC/JU	Ok
15	P4460-03	WB-303-BOT	BF140063.D	26 Oct 2024 16:48		RC/JU	Ok
16	P4471-02	B-180-SB02	BF140064.D	26 Oct 2024 17:16		RC/JU	Ok
17	P4508-12	BP-F22	BF140065.D	26 Oct 2024 17:44		RC/JU	Ok
18	P4508-04	TP-3	BF140066.D	26 Oct 2024 18:12		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF102624

Review By	yogesh	Review On	10/29/2024 1:34:36 AM		
Supervise By	mohammad	Supervise On	10/29/2024 1:38:32 AM		
SubDirectory	BF102624	HP Acquire Method	BNA_F	HP Processing Method	bf101824
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12323,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4508-08	BP-F23	BF140067.D	26 Oct 2024 18:40		RC/JU	Ok
20	P4460-02	WB-303-TOP	BF140068.D	26 Oct 2024 19:09		RC/JU	Ok,M
21	P4547-01	BP-F-21	BF140069.D	26 Oct 2024 19:37		RC/JU	Ok
22	P4508-01	TP-3	BF140070.D	26 Oct 2024 20:05		RC/JU	Ok
23	P4545-01	VNJ-215	BF140071.D	26 Oct 2024 20:33		RC/JU	Ok,M
24	P4509-01	AU-06-10232024	BF140072.D	26 Oct 2024 21:01		RC/JU	Ok,M
25	P4531-01	OR-03-102424	BF140073.D	26 Oct 2024 21:29		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF102924

Review By	Jagrut	Review On	10/30/2024 10:58:28 AM
Supervise By	mohammad	Supervise On	11/4/2024 1:49:36 AM
SubDirectory	BF102924	HP Acquire Method	BNA_F
		HP Processing Method	bf101824

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12323,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140099.D	29 Oct 2024 09:09		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140100.D	29 Oct 2024 09:37		RC/JU	Ok,M
3	PB164461BL	PB164461BL	BF140101.D	29 Oct 2024 10:05		RC/JU	Ok
4	PB164461BS	PB164461BS	BF140102.D	29 Oct 2024 10:33		RC/JU	Ok,M
5	P4495-12	PT-TRIAZINE-SOIL	BF140103.D	29 Oct 2024 11:01		RC/JU	Ok
6	P4561-05	BP-F-18	BF140104.D	29 Oct 2024 11:34		RC/JU	Ok
7	P4561-05MS	BP-F-18MS	BF140105.D	29 Oct 2024 12:02		RC/JU	Ok,M
8	P4561-05MSD	BP-F-18MSD	BF140106.D	29 Oct 2024 12:30		RC/JU	Ok,M
9	P4574-07	CONCRETE-1	BF140107.D	29 Oct 2024 12:58		RC/JU	Ok
10	P4574-09	CONCRETE-2	BF140108.D	29 Oct 2024 13:24		RC/JU	Ok
11	P4574-11	CONCRETE-3	BF140109.D	29 Oct 2024 13:50		RC/JU	Ok
12	P4578-05	WB-305-BOT	BF140110.D	29 Oct 2024 14:16		RC/JU	Ok
13	P4578-07	WB-305-BOT-1	BF140111.D	29 Oct 2024 14:42		RC/JU	Ok
14	DFTPP	DFTPP	BF140112.D	29 Oct 2024 15:35		RC/JU	Ok
15	SSTDCCC040	SSTDCCC040	BF140113.D	29 Oct 2024 16:01		RC/JU	Ok,M
16	PB164497BL	PB164497BL	BF140114.D	29 Oct 2024 16:26		RC/JU	Ok
17	PB164497BS	PB164497BS	BF140115.D	29 Oct 2024 16:52		RC/JU	Ok,M
18	P4594-05	BP-F17	BF140116.D	29 Oct 2024 17:18		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF102924

Review By	Jagrut	Review On	10/30/2024 10:58:28 AM		
Supervise By	mohammad	Supervise On	11/4/2024 1:49:36 AM		
SubDirectory	BF102924	HP Acquire Method	BNA_F	HP Processing Method	bf101824
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12323,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4594-05MS	BP-F17MS	BF140117.D	29 Oct 2024 17:45		RC/JU	Ok,M
20	P4594-05MSD	BP-F17MSD	BF140118.D	29 Oct 2024 18:11		RC/JU	Ok,M
21	P4597-07	BLUE-2-1	BF140119.D	29 Oct 2024 18:37		RC/JU	Ok,M
22	P4594-17	BP-F15	BF140120.D	29 Oct 2024 19:03		RC/JU	Ok
23	P4598-05	BP-F11	BF140121.D	29 Oct 2024 19:29		RC/JU	Ok
24	P4567-05	WC-2	BF140122.D	29 Oct 2024 19:55		RC/JU	Ok
25	P4597-01	RED-1-1	BF140123.D	29 Oct 2024 20:21		RC/JU	Ok,M
26	P4574-01	GRAVEL-1	BF140124.D	29 Oct 2024 20:47		RC/JU	Ok
27	P4578-04	WB-305-TOP-1	BF140125.D	29 Oct 2024 21:13		RC/JU	Ok,M
28	P4561-01	BP-F-19	BF140126.D	29 Oct 2024 21:39		RC/JU	Ok
29	P4567-09	WC-3	BF140127.D	29 Oct 2024 22:05		RC/JU	Ok
30	P4598-01	BP-F12	BF140128.D	29 Oct 2024 22:31		RC/JU	Ok
31	P4597-04	RED-1-2	BF140129.D	29 Oct 2024 22:57		RC/JU	Ok
32	P4597-10	BLUE-2-2	BF140130.D	29 Oct 2024 23:24		RC/JU	Ok
33	P4594-13	TP-5	BF140131.D	29 Oct 2024 23:49		RC/JU	Ok
34	DFTPP	DFTPP	BF140132.D	30 Oct 2024 00:41		RC/JU	Ok
35	SSTDCCC040	SSTDCCC040	BF140133.D	30 Oct 2024 01:11		RC/JU	Ok,M
36	PB164500BL	PB164500BL	BF140134.D	30 Oct 2024 02:04		RC/JU	Ok
37	PB164478TB	PB164478TB	BF140135.D	30 Oct 2024 02:30		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF102924

Review By	Jagrut	Review On	10/30/2024 10:58:28 AM		
Supervise By	mohammad	Supervise On	11/4/2024 1:49:36 AM		
SubDirectory	BF102924	HP Acquire Method	BNA_F	HP Processing Method	bf101824
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12323,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

38	PB164500BS	PB164500BS	BF140136.D	30 Oct 2024 02:56		RC/JU	Ok,M
39	P4597-03	RED-1-1	BF140137.D	30 Oct 2024 03:22		RC/JU	Ok
40	P4597-06	RED-1-2	BF140138.D	30 Oct 2024 03:48		RC/JU	Ok
41	P4594-04	TP-4	BF140139.D	30 Oct 2024 04:14		RC/JU	Ok
42	P4594-04MS	TP-4MS	BF140140.D	30 Oct 2024 04:40		RC/JU	Ok,M
43	P4594-04MSD	TP-4MSD	BF140141.D	30 Oct 2024 05:06		RC/JU	Ok,M
44	P4594-08	BP-F17	BF140142.D	30 Oct 2024 05:32		RC/JU	Ok
45	P4594-12	BP-F16	BF140143.D	30 Oct 2024 05:58		RC/JU	Ok
46	P4594-16	TP-5	BF140144.D	30 Oct 2024 06:24		RC/JU	Ok
47	P4594-20	BP-F15	BF140145.D	30 Oct 2024 06:50		RC/JU	Ok
48	P4597-09	BLUE-2-1	BF140146.D	30 Oct 2024 07:16		RC/JU	Ok
49	P4597-12	BLUE-2-2	BF140147.D	30 Oct 2024 07:42		RC/JU	Ok
50	P4598-04	BP-F12	BF140148.D	30 Oct 2024 08:08		RC/JU	Ok
51	P4598-08	BP-F11	BF140149.D	30 Oct 2024 08:34		RC/JU	Ok
52	P4598-12	TP-8	BF140150.D	30 Oct 2024 09:00		RC/JU	Ok

M : Manual Integration