

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/26/2024

Lab File ID: BF140050.D Time Analyzed: 10:38

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	128827	6.892	465620	8.18	252523	9.93
UPPER LIMIT	257654	7.392	931240	8.675	505046	10.428
LOWER LIMIT	64413.5	6.392	232810	7.675	126262	9.428
EPA SAMPLE NO.						
01 PB164401BL	125232	6.89	486243	8.17	274404	9.92
02 PB164401BS	133953	6.89	523078	8.17	286545	9.93
03 BP-F-20MS	141513	6.89	556548	8.17	317233	9.93
04 BP-F-20MSD	142054	6.89	561979	8.17	322249	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.